

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021622\
 Data File : BF126912.D
 Acq On : 16 Feb 2022 14:39
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF021622

Quant Time: Feb 16 17:05:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:55:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.592	0.578	2.4	106	0.00
3	Pyridine	1.553	1.545	0.5	105	0.00
4	n-Nitrosodimethylamine	0.761	0.753	1.1	102	0.00
5 S	2-Fluorophenol	1.294	1.244	3.9	103	0.00
6	Aniline	2.069	2.059	0.5	103	0.00
7 S	Phenol-d6	1.746	1.707	2.2	104	0.00
8	2-Chlorophenol	1.388	1.371	1.2	104	0.00
9	Benzaldehyde	1.063	0.963	9.4	107	0.00
10 C	Phenol	1.764	1.718	2.6	105	0.00
11	bis(2-Chloroethyl)ether	1.384	1.349	2.5	104	0.00
12	1,3-Dichlorobenzene	1.498	1.456	2.8	105	0.00
13 C	1,4-Dichlorobenzene	1.518	1.469	3.2	104	0.00
14	1,2-Dichlorobenzene	1.448	1.396	3.6	103	0.00
15	Benzyl Alcohol	1.471	1.461	0.7	103	0.00
16	2,2'-oxybis(1-Chloropropane	2.179	2.109	3.2	102	0.00
17	2-Methylphenol	1.201	1.191	0.8	104	0.00
18	Hexachloroethane	0.582	0.570	2.1	105	0.00
19 P	n-Nitroso-di-n-propylamine	1.125	1.120	0.4	106	0.00
20	3+4-Methylphenols	1.560	1.543	1.1	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.527	0.493	6.5	103	0.00
23 S	Nitrobenzene-d5	0.443	0.419	5.4	103	0.00
24	Nitrobenzene	0.419	0.400	4.5	104	0.00
25	Isophorone	0.733	0.704	4.0	104	0.00
26 C	2-Nitrophenol	0.178	0.180	-1.1	107	0.00
27	2,4-Dimethylphenol	0.293	0.280	4.4	105	0.00
28	bis(2-Chloroethoxy)methane	0.445	0.427	4.0	105	0.00
29 C	2,4-Dichlorophenol	0.275	0.266	3.3	105	0.00
30	1,2,4-Trichlorobenzene	0.308	0.293	4.9	105	0.00
31	Naphthalene	1.044	0.986	5.6	104	0.00
32	Benzoic acid	0.208	0.220	-5.8	105	0.00
33	4-Chloroaniline	0.411	0.391	4.9	103	0.00
34 C	Hexachlorobutadiene	0.180	0.172	4.4	105	0.00
35	Caprolactam	0.096	0.093	3.1	103	0.00
36 C	4-Chloro-3-methylphenol	0.352	0.343	2.6	103	0.00
37	2-Methylnaphthalene	0.677	0.638	5.8	103	0.00
38	1-Methylnaphthalene	0.649	0.621	4.3	105	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.526	0.507	3.6	105	0.00
41 P	Hexachlorocyclopentadiene	0.286	0.295	-3.1	106	0.00
42 S	2,4,6-Tribromophenol	0.230	0.223	3.0	101	0.00
43 C	2,4,6-Trichlorophenol	0.386	0.372	3.6	102	0.00
44	2,4,5-Trichlorophenol	0.406	0.394	3.0	103	0.00
45 S	2-Fluorobiphenyl	1.359	1.277	6.0	103	0.00
46	1,1'-Biphenyl	1.585	1.503	5.2	103	0.00
47	2-Chloronaphthalene	1.175	1.115	5.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021622\
 Data File : BF126912.D
 Acq On : 16 Feb 2022 14:39
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF021622

Quant Time: Feb 16 17:05:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:55:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.464	0.459	1.1	105	0.00
49	Acenaphthylene	1.909	1.831	4.1	103	0.00
50	Dimethylphthalate	1.430	1.378	3.6	104	0.00
51	2,6-Dinitrotoluene	0.291	0.285	2.1	103	0.00
52 C	Acenaphthene	1.241	1.211	2.4	104	0.00
53	3-Nitroaniline	0.356	0.347	2.5	102	0.00
54 P	2,4-Dinitrophenol	0.154	0.166	-7.8	105	0.00
55	Dibenzofuran	1.707	1.631	4.5	103	0.00
56 P	4-Nitrophenol	0.263	0.264	-0.4	103	0.00
57	2,4-Dinitrotoluene	0.393	0.386	1.8	102	0.00
58	Fluorene	1.330	1.251	5.9	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.322	0.314	2.5	103	0.00
60	Diethylphthalate	1.494	1.427	4.5	102	0.00
61	4-Chlorophenyl-phenylether	0.627	0.598	4.6	102	0.00
62	4-Nitroaniline	0.351	0.345	1.7	102	0.00
63	Azobenzene	1.663	1.577	5.2	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.126	-6.8	105	0.00
66 c	n-Nitrosodiphenylamine	0.656	0.623	5.0	101	0.00
67	4-Bromophenyl-phenylether	0.223	0.213	4.5	102	0.00
68	Hexachlorobenzene	0.241	0.230	4.6	103	0.00
69	Atrazine	0.203	0.198	2.5	105	0.00
70 C	Pentachlorophenol	0.131	0.135	-3.1	101	0.00
71	Phenanthrene	1.119	1.050	6.2	102	0.00
72	Anthracene	1.114	1.054	5.4	101	0.00
73	Carbazole	1.022	0.959	6.2	101	0.00
74	Di-n-butylphthalate	1.314	1.263	3.9	102	0.00
75 C	Fluoranthene	1.063	0.999	6.0	101	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzydine	0.544	0.502	7.7	93	0.00
78	Pyrene	1.738	1.705	1.9	102	0.00
79 S	Terphenyl-d14	1.361	1.310	3.7	103	0.00
80	Butylbenzylphthalate	0.823	0.816	0.9	102	0.00
81	Benzo(a)anthracene	1.348	1.314	2.5	103	0.00
82	3,3'-Dichlorobenzidine	0.425	0.411	3.3	100	0.00
83	Chrysene	1.268	1.230	3.0	103	0.00
84	Bis(2-ethylhexyl)phthalate	1.080	1.076	0.4	104	0.00
85 c	Di-n-octyl phthalate	1.555	1.540	1.0	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	1.311	1.357	-3.5	101	0.00
88	Benzo(b)fluoranthene	1.343	1.308	2.6	103	0.00
89	Benzo(k)fluoranthene	1.296	1.337	-3.2	106	0.00
90 C	Benzo(a)pyrene	1.049	1.062	-1.2	104	0.00
91	Dibenzo(a,h)anthracene	1.085	1.120	-3.2	102	0.00
92	Benzo(g,h,i)perylene	1.069	1.098	-2.7	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021622\
Data File : BF126912.D
Acq On : 16 Feb 2022 14:39
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF021622

Quant Time: Feb 16 17:05:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 16 14:55:48 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0