

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021623\
 Data File : BF132440.D
 Acq On : 16 Feb 2023 18:45
 Operator : CG\JU
 Sample : 01537-02RE
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEST-PIT-1-(TP1)-COMP-1RE

Quant Time: Feb 17 04:44:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 17 03:08:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.031	152	196155	20.000	ng	0.00
21) Naphthalene-d8	8.313	136	737430	20.000	ng	0.00
39) Acenaphthene-d10	10.077	164	194227	20.000	ng	0.00
64) Phenanthrene-d10	11.572	188	600405	20.000	ng	0.00
76) Chrysene-d12	14.230	240	111329	20.000	ng	#-0.01
86) Perylene-d12	15.789	264	2753	20.000	ng	#-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.672	112	1633516	133.440	ng	0.02
7) Phenol-d6	6.648	99	1348215	84.345	ng	0.00
23) Nitrobenzene-d5	7.595	82	1543336	100.560	ng	0.00
42) 2,4,6-Tribromophenol	10.872	330	569220	293.733	ng	0.00
45) 2-Fluorobiphenyl	9.395	172	2225665	176.963	ng	0.00
79) Terphenyl-d14	13.171	244	2142579	350.499	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.777	163	31125	2.176	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021623\
Data File : BF132440.D
Acq On : 16 Feb 2023 18:45
Operator : CG\JU
Sample : 01537-02RE
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEST-PIT-1-(TP1)-COMP-1RE

Quant Time: Feb 17 04:44:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 03:08:40 2023
Response via : Initial Calibration

