


Data Path : Z:\HPCHEM1\BNA F\DATA\BF021716\  
 Data File : BF085081.D  
 Acq On : 17 Feb 2016 9:43  
 Operator : SJ/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF021716.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Feb 17 16:42:47 2016



AutoFind: Scans 890, 891, 892; Background Corrected with Scan 887

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 41.3      | 47803   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.3       | 594     | PASS             |
| 69          | 198          | 0.00         | 100          | 38.4      | 44449   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 193     | PASS             |
| 127         | 198          | 10           | 80           | 50.7      | 58619   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 451     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 115698  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 7868    | PASS             |
| 275         | 198          | 10           | 60           | 26.7      | 30889   | PASS             |
| 365         | 198          | 1            | 100          | 3.3       | 3767    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.7      | 17094   | PASS             |
| 442         | 198          | 50           | 100          | 94.3      | 109108  | PASS             |
| 443         | 442          | 15           | 24           | 19.7      | 21541   | PASS             |