

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021722\
 Data File : BF126922.D
 Acq On : 17 Feb 2022 12:33
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 17 17:07:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:55:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	40.576	-1.4	103	0.00
3	Pyridine	40.000	40.481	-1.2	100	0.00
4	n-Nitrosodimethylamine	40.000	39.201	2.0	95	0.00
5 S	2-Fluorophenol	80.000	78.551	1.8	98	0.00
6	Aniline	40.000	39.530	1.2	96	0.00
7 S	Phenol-d6	80.000	78.273	2.2	97	0.00
8	2-Chlorophenol	40.000	39.429	1.4	98	0.00
9	Benzaldehyde	40.000	35.097	12.3	97	0.00
10 C	Phenol	40.000	39.599	1.0	100	0.00
11	bis(2-Chloroethyl)ether	40.000	38.777	3.1	96	0.00
12	1,3-Dichlorobenzene	40.000	39.382	1.5	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.236	1.9	99	0.00
14	1,2-Dichlorobenzene	40.000	38.612	3.5	97	0.00
15	Benzyl Alcohol	40.000	39.221	1.9	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.798	5.5	93	0.00
17	2-Methylphenol	40.000	39.052	2.4	96	0.00
18	Hexachloroethane	40.000	39.624	0.9	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.437	3.9	96	0.00
20	3+4-Methylphenols	40.000	38.841	2.9	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	37.468	6.3	95	0.00
23 S	Nitrobenzene-d5	80.000	76.231	4.7	96	0.00
24	Nitrobenzene	40.000	38.308	4.2	96	0.00
25	Isophorone	40.000	37.507	6.2	93	0.00
26 C	2-Nitrophenol	40.000	39.625	0.9	97	0.00
27	2,4-Dimethylphenol	40.000	37.820	5.4	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.884	5.3	95	0.00
29 C	2,4-Dichlorophenol	40.000	38.284	4.3	95	0.00
30	1,2,4-Trichlorobenzene	40.000	38.199	4.5	96	0.00
31	Naphthalene	40.000	38.010	5.0	97	0.00
32	Benzoic acid	40.000	39.880	0.3	91	0.00
33	4-Chloroaniline	40.000	38.092	4.8	95	0.00
34 C	Hexachlorobutadiene	40.000	37.984	5.0	95	0.00
35	Caprolactam	40.000	38.825	2.9	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.673	3.3	94	0.00
37	2-Methylnaphthalene	40.000	37.749	5.6	95	0.00
38	1-Methylnaphthalene	40.000	37.918	5.2	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.436	3.9	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.457	-1.1	95	0.00
42 S	2,4,6-Tribromophenol	80.000	78.971	1.3	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.299	4.3	94	0.00
44	2,4,5-Trichlorophenol	40.000	38.388	4.0	94	0.00
45 S	2-Fluorobiphenyl	80.000	75.563	5.5	96	0.00
46	1,1'-Biphenyl	40.000	38.118	4.7	95	0.00
47	2-Chloronaphthalene	40.000	38.400	4.0	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021722\
 Data File : BF126922.D
 Acq On : 17 Feb 2022 12:33
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 17 17:07:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:55:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.158	2.1	96	0.00
49	Acenaphthylene	40.000	38.663	3.3	95	0.00
50	Dimethylphthalate	40.000	38.052	4.9	94	0.00
51	2,6-Dinitrotoluene	40.000	38.978	2.6	94	0.00
52 C	Acenaphthene	40.000	38.692	3.3	95	0.00
53	3-Nitroaniline	40.000	38.894	2.8	94	0.00
54 P	2,4-Dinitrophenol	40.000	41.065	-2.7	93	0.00
55	Dibenzofuran	40.000	38.115	4.7	95	0.00
56 P	4-Nitrophenol	40.000	40.255	-0.6	95	0.00
57	2,4-Dinitrotoluene	40.000	39.924	0.2	96	0.00
58	Fluorene	40.000	38.209	4.5	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.275	4.3	93	0.00
60	Diethylphthalate	40.000	38.269	4.3	94	0.00
61	4-Chlorophenyl-phenylether	40.000	38.518	3.7	95	0.00
62	4-Nitroaniline	40.000	40.194	-0.5	96	0.00
63	Azobenzene	40.000	38.076	4.8	94	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.359	-5.9	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.389	4.0	95	0.00
67	4-Bromophenyl-phenylether	40.000	39.544	1.1	98	0.00
68	Hexachlorobenzene	40.000	39.308	1.7	98	0.00
69	Atrazine	40.000	38.952	2.6	97	0.00
70 C	Pentachlorophenol	40.000	41.635	-4.1	95	0.00
71	Phenanthrene	40.000	38.658	3.4	98	0.00
72	Anthracene	40.000	38.494	3.8	96	0.00
73	Carbazole	40.000	38.963	2.6	97	0.00
74	Di-n-butylphthalate	40.000	39.354	1.6	97	0.00
75 C	Fluoranthene	40.000	39.857	0.4	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	40.121	-0.3	101	0.00
78	Pyrene	40.000	38.229	4.4	99	0.00
79 S	Terphenyl-d14	80.000	77.139	3.6	102	0.00
80	Butylbenzylphthalate	40.000	40.019	-0.0	103	0.00
81	Benzo(a)anthracene	40.000	39.046	2.4	103	0.00
82	3,3'-Dichlorobenzidine	40.000	39.651	0.9	102	0.00
83	Chrysene	40.000	38.719	3.2	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.952	0.1	104	0.00
85 c	Di-n-octyl phthalate	40.000	39.107	2.2	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.615	-4.0	98	0.00
88	Benzo(b)fluoranthene	40.000	39.239	1.9	100	0.00
89	Benzo(k)fluoranthene	40.000	40.228	-0.6	99	0.00
90 C	Benzo(a)pyrene	40.000	39.182	2.0	96	0.00
91	Dibenzo(a,h)anthracene	40.000	40.871	-2.2	97	0.00
92	Benzo(g,h,i)perylene	40.000	42.158	-5.4	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021722\
Data File : BF126922.D
Acq On : 17 Feb 2022 12:33
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 17 17:07:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 16 14:55:48 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0