


Data Path : Z:\HPCHEM1\BNA_F\Data\BF021815\
 Data File : BF077353.D
 Acq On : 18 Feb 2015 15:32
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021715.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Feb 18 09:51:00 2015



AutoFind: Scans 1087, 1088, 1089; Background Corrected with Scan 1084

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.3	6538	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.7	6272	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.8	8191	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	15817	PASS
199	198	5	9	6.3	991	PASS
275	198	10	60	26.4	4181	PASS
365	198	1	100	3.8	598	PASS
441	442	0.01	24	15.7	2376	PASS
442	198	50	100	95.5	15113	PASS
443	442	15	24	19.9	3000	PASS