

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF021819\
 Data File : BF112603.D
 Acq On : 18 Feb 2019 16:13
 Operator : JU/SJ
 Sample : K1518-15
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CABLE-BRIDGE-FOUNDATION-D-8

Manual Integrations
 APPROVED

Sohil
 2/19/2019 9:38:42 AM

Quant Time: Feb 19 00:07:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	73153	20.00	ng	-0.02
21) Naphthalene-d8	8.10	136	290760	20.00	ng	-0.02
39) Acenaphthene-d10	9.86	164	144305	20.00	ng	-0.02
64) Phenanthrene-d10	11.35	188	259446	20.00	ng	-0.02
76) Chrysene-d12	14.00	240	183975	20.00	ng	-0.01
87) Perylene-d12	15.47	264	171083	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	325806	94.60	ng	-0.02
7) Phenol-d6	6.47	99	417485	87.24	ng	-0.02
23) Nitrobenzene-d5	7.39	82	266423	52.80	ng	-0.02
42) 2,4,6-Tribromophenol	10.66	330	135365	80.59	ng	-0.02
45) 2-Fluorobiphenyl	9.17	172	528189	51.77	ng	-0.03
79) Terphenyl-d14	12.93	244	500193	51.21	ng	-0.02
Target Compounds						
10) Phenol	6.49	94	12185	2.321	ng	Qvalue # 65
50) Dimethylphthalate	9.57	163	48284	4.307	ng	100
75) Fluoranthene	12.57	202	96696	6.684	ng	96
78) Pyrene	12.80	202	93608	7.349	ng	98
81) Benzo(a)anthracene	13.99	228	26075	2.411	ng	98
83) Chrysene	14.02	228	45153	4.374	ng	98
84) Bis(2-ethylhexyl)phthalate	13.96	149	739334	84.518	ng	97
86) Indeno(1,2,3-cd)pyrene	16.98	276	22381	2.736	ng	95
88) Benzo(b)fluoranthene	15.04	252	63915m	6.247	ng	
89) Benzo(k)fluoranthene	15.07	252	20313m	2.073	ng	
90) Benzo(a)pyrene	15.42	252	22465	2.440	ng	# 89
92) Benzo(a,h,i)perylene	17.43	276	22070	3.153	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF021819\
Data File : BF112603.D
Acq On : 18 Feb 2019 16:13
Operator : JU/SJ
Sample : K1518-15
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-8

Manual Integrations
APPROVED

Sohil
2/19/2019 9:38:42 AM

Quant Time: Feb 19 00:07:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration

