




Data Path : Z:\HPCHEM1\BNA F\DATA\BF021918\
Data File : BF103095.D
Acq On : 19 Feb 2018 21:24
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 20 05:54:39 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 16 14:34:49 2018
Response via : Initial Calibration



TIC: BF103095.D

(69) Pentachlorophenol (C)




11.210min (-0.000) 28910.32ng

response 61790

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	59.56
264.00	61.90	61.56
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF021918\
 Data File : BF103095.D
 Acq On : 19 Feb 2018 21:24
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 20 05:54:39 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 16 14:34:49 2018
 Response via : Initial Calibration



TIC: BF103095.D

(76) Benzidine

12.739min (-0.006) 0.00ng

response 481958

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	15.36
183.00	11.60	11.46
0.00	0.00	0.00