

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022015\
 Data File : BF077394.D
 Acq On : 20 Feb 2015 21:05
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Feb 21 01:38:14 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 20 23:55:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.508	0.523	-3.0	121	0.00
3	Pyridine	1.454	1.160	20.2	87	0.00
4	n-Nitrosodimethylamine	0.632	0.653	-3.3	124	0.00
5 S	2-Fluorophenol	1.198	1.198	0.0	115	0.00
6	Aniline	2.129	2.127	0.1	112	0.00
7 S	Phenol-d6	1.540	1.482	3.8	108	0.00
8	2-Chlorophenol	1.413	1.420	-0.5	115	0.00
9	Benzaldehyde	0.935	0.875	6.4	110	-0.01
10 C	Phenol	1.828	1.840	-0.7	110	0.00
11	bis(2-Chloroethyl)ether	1.313	1.246	5.1	110	-0.01
12	1,3-Dichlorobenzene	1.539	1.542	-0.2	114	0.00
13 C	1,4-Dichlorobenzene	1.566	1.565	0.1	113	0.00
14	1,2-Dichlorobenzene	1.489	1.511	-1.5	114	0.00
15	Benzyl Alcohol	1.171	1.128	3.7	109	0.00
16	2,2'-oxybis(1-Chloropropane	1.877	1.860	0.9	111	-0.01
17	2-Methylphenol	1.105	1.141	-3.3	111	0.00
18	Hexachloroethane	0.523	0.521	0.4	113	-0.01
19 P	n-Nitroso-di-n-propylamine	0.978	1.035	-5.8	117	0.00
20	3+4-Methylphenols	1.455	1.437	1.2	110	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
22	Acetophenone	0.470	0.469	0.2	113	0.00
23 S	Nitrobenzene-d5	0.322	0.324	-0.6	108	-0.01
24	Nitrobenzene	0.338	0.343	-1.5	110	-0.01
25	Isophorone	0.638	0.652	-2.2	106	-0.01
26 C	2-Nitrophenol	0.139	0.157	-12.9	113	0.00
27	2,4-Dimethylphenol	0.310	0.324	-4.5	113	0.00
28	bis(2-Chloroethoxy)methane	0.403	0.438	-8.7	117	0.00
29 C	2,4-Dichlorophenol	0.291	0.310	-6.5	114	-0.01
30	1,2,4-Trichlorobenzene	0.320	0.337	-5.3	114	0.00
31	Naphthalene	1.000	1.023	-2.3	114	0.00
32	Benzoic acid	0.151	0.176	-16.6	118	0.00
33	4-Chloroaniline	0.445	0.439	1.3	104	-0.01
34 C	Hexachlorobutadiene	0.183	0.195	-6.6	116	0.00
35	Caprolactam	0.089	0.094	-5.6	111	0.00
36 C	4-Chloro-3-methylphenol	0.293	0.315	-7.5	112	0.00
37	2-Methylnaphthalene	0.710	0.719	-1.3	106	-0.01
38 I	Acenaphthene-d10	1.000	1.000	0.0	112	-0.01
39	1,2,4,5-Tetrachlorobenzene	0.574	0.570	0.7	107	-0.01
40 P	Hexachlorocyclopentadiene	0.308	0.312	-1.3	103	-0.01
41 S	2,4,6-Tribromophenol	0.193	0.202	-4.7	109	0.00
42 C	2,4,6-Trichlorophenol	0.380	0.405	-6.6	110	0.00
43	2,4,5-Trichlorophenol	0.406	0.408	-0.5	106	0.00
44 S	2-Fluorobiphenyl	1.283	1.313	-2.3	114	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022015\
 Data File : BF077394.D
 Acq On : 20 Feb 2015 21:05
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Feb 21 01:38:14 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 20 23:55:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.515	1.484	2.0	106	-0.01
46	2-Chloronaphthalene	1.188	1.165	1.9	109	0.00
47	2-Nitroaniline	0.293	0.310	-5.8	113	0.00
48	Acenaphthylene	1.881	1.880	0.1	110	0.00
49	Dimethylphthalate	1.406	1.366	2.8	107	0.00
50	2,6-Dinitrotoluene	0.282	0.290	-2.8	102	-0.01
51 C	Acenaphthene	1.123	1.143	-1.8	110	0.00
52	3-Nitroaniline	0.325	0.358	-10.2	113	0.00
53 P	2,4-Dinitrophenol	0.086	0.079	8.1	98	-0.01
54	Dibenzofuran	1.733	1.766	-1.9	111	0.00
55 P	4-Nitrophenol	0.261	0.258	1.1	99	-0.01
56	2,4-Dinitrotoluene	0.381	0.399	-4.7	103	0.00
57	Fluorene	1.386	1.388	-0.1	108	0.00
58	2,3,4,6-Tetrachlorophenol	0.327	0.342	-4.6	107	0.00
59	Diethylphthalate	1.386	1.383	0.2	108	0.00
60	4-Chlorophenyl-phenylether	0.668	0.656	1.8	106	0.00
61	4-Nitroaniline	0.312	0.330	-5.8	111	0.00
62	Azobenzene	1.315	1.347	-2.4	108	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
64	4,6-Dinitro-2-methylphenol	0.075	0.069	8.0	93	-0.01
65 c	n-Nitrosodiphenylamine	0.664	0.678	-2.1	104	-0.01
66	4-Bromophenyl-phenylether	0.234	0.237	-1.3	108	0.00
67	Hexachlorobenzene	0.251	0.255	-1.6	110	0.00
68	Atrazine	0.216	0.194	10.2	101	0.00
69 C	Pentachlorophenol	0.132	0.138	-4.5	104	0.00
70	Phenanthrene	1.159	1.128	2.7	104	-0.01
71	Anthracene	1.150	1.147	0.3	108	0.00
72	Carbazole	1.094	1.151	-5.2	107	0.00
73	Di-n-butylphthalate	1.284	1.271	1.0	99	0.00
74 C	Fluoranthene	1.266	1.287	-1.7	106	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	98	0.01
76	Benzidine	0.676	0.613	9.3	96	0.00
77	Pyrene	1.223	1.244	-1.7	101	-0.01
78 S	Terphenyl-d14	0.856	0.872	-1.9	100	0.00
79	Butylbenzylphthalate	0.503	0.554	-10.1	103	0.00
80	Benzo(a)anthracene	1.172	1.170	0.2	100	0.01
81	3,3'-Dichlorobenzidine	0.430	0.456	-6.0	102	0.01
82	Chrysene	1.101	1.094	0.6	101	0.01
83	Bis(2-ethylhexyl)phthalate	0.807	0.827	-2.5	99	0.01
84 c	Di-n-octyl phthalate	1.348	1.355	-0.5	94	0.05
85	Indeno(1,2,3-cd)pyrene	1.255	1.290	-2.8	101	0.09
86 I	Perylene-d12	1.000	1.000	0.0	99	0.07
87	Benzo(b)fluoranthene	1.163	1.174	-0.9	94	0.06

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022015\
Data File : BF077394.D
Acq On : 20 Feb 2015 21:05
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Feb 21 01:38:14 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 20 23:55:25 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.140	1.163	-2.0	108	0.07
89 C	Benzo(a)pyrene	1.108	1.137	-2.6	98	0.08
90	Dibenzo(a,h)anthracene	1.056	1.093	-3.5	100	0.09
91	Benzo(g,h,i)perylene	1.066	1.113	-4.4	101	0.09

(#) = Out of Range

SPCC's out = 0 CCC's out = 0