

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022017\
 Data File : BF093072.D
 Acq On : 21 Feb 2017 13:08
 Operator : SJ/MA
 Sample : I1910-01RE
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HARMONBLDG30-RW-1RE

Manual Integrations
 APPROVED

mohammad
 2/21/2017 3:58:05 PM

Quant Time: Feb 21 14:38:33 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 17 17:31:55 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	137424m	20.00	ng	-0.01
21) Naphthalene-d8	8.14	136	420272m	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	175388m	20.00	ng	0.02
63) Phenanthrene-d10	11.40	188	290093m	20.00	ng	0.02
75) Chrysene-d12	14.01	240	276103m	20.00	ng	-0.01
86) Perylene-d12	15.44	264	288563m	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	496117	61.94	ng	-0.01
7) Phenol-d6	6.48	99	384974	37.35	ng	-0.01
23) Nitrobenzene-d5	7.42	82	877528	116.28	ng	0.00
41) 2,4,6-Tribromophenol	10.70	330	187600m	147.89	ng	0.02
44) 2-Fluorobiphenyl	9.23	172	1069059m	110.43	ng	0.01
78) Terphenyl-d14	12.97	244	845804m	71.98	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.86	142	363909m	26.60	ng	Qvalue
51) Acenaphthene	9.95	154	94816m	9.24	ng	
70) Phenanthrene	11.42	178	444143m	26.72	ng	
74) Fluoranthene	12.60	202	41377m	2.41	ng	
77) Pyrene	12.83	202	161954m	7.84	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022017\
Data File : BF093072.D
Acq On : 21 Feb 2017 13:08
Operator : SJ/MA
Sample : I1910-01RE
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HARMONBLDG30-RW-1RE

Manual Integrations
APPROVED

mohammad
2/21/2017 3:58:05 PM

Quant Time: Feb 21 14:38:33 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 17:31:55 2017
Response via : Initial Calibration

