




Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG022020\
 Data File : BG044522.D
 Acq On : 20 Feb 2020 15:59
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 21 08:11:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 10 16:11:50 2020
 Response via : Initial Calibration



TIC: BG044522.D

(77) Benzidine

19.500min (-0.022) 0.00ng

response 451986

Ion	Exp%	Act%
184.00	100	100
185.00	16.10	14.79
183.00	13.50	12.59
0.00	0.00	0.00