

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022020\
 Data File : BF118966.D
 Acq On : 20 Feb 2020 18:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 21 02:10:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 20 17:40:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.533	0.510	4.3	108	0.01
3	Pyridine	1.455	1.399	3.8	109	0.00
4	n-Nitrosodimethylamine	0.441	0.428	2.9	109	0.00
5 S	2-Fluorophenol	1.197	1.147	4.2	106	0.00
6	Aniline	1.796	1.740	3.1	106	0.00
7 S	Phenol-d6	1.455	1.402	3.6	106	0.00
8	2-Chlorophenol	1.292	1.239	4.1	106	0.00
9	Benzaldehyde	0.972	0.960	1.2	110	0.00
10 C	Phenol	1.574	1.528	2.9	108	0.00
11	bis(2-Chloroethyl)ether	1.204	1.153	4.2	107	0.00
12	1,3-Dichlorobenzene	1.548	1.482	4.3	107	0.00
13 C	1,4-Dichlorobenzene	1.549	1.503	3.0	108	0.00
14	1,2-Dichlorobenzene	1.413	1.379	2.4	108	0.00
15	Benzyl Alcohol	1.052	1.020	3.0	106	0.00
16	2,2'-oxybis(1-Chloropropane	1.043	1.010	3.2	109	0.00
17	2-Methylphenol	1.047	1.001	4.4	107	0.00
18	Hexachloroethane	0.554	0.530	4.3	107	0.00
19 P	n-Nitroso-di-n-propylamine	0.848	0.801	5.5	107	0.00
20	3+4-Methylphenols	1.283	1.188	7.4	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
22	Acetophenone	0.463	0.430	7.1	106	0.00
23 S	Nitrobenzene-d5	0.379	0.356	6.1	107	0.00
24	Nitrobenzene	0.375	0.352	6.1	108	0.00
25	Isophorone	0.658	0.609	7.4	107	0.00
26 C	2-Nitrophenol	0.198	0.186	6.1	107	0.00
27	2,4-Dimethylphenol	0.269	0.250	7.1	106	0.00
28	bis(2-Chloroethoxy)methane	0.428	0.399	6.8	107	0.00
29 C	2,4-Dichlorophenol	0.317	0.300	5.4	107	0.00
30	1,2,4-Trichlorobenzene	0.376	0.355	5.6	107	0.00
31	Naphthalene	1.037	0.972	6.3	106	0.00
32	Benzoic acid	0.184	0.168	8.7	106	0.00
33	4-Chloroaniline	0.446	0.414	7.2	105	0.00
34 C	Hexachlorobutadiene	0.234	0.221	5.6	108	0.00
35	Caprolactam	0.098	0.088	10.2	101	0.00
36 C	4-Chloro-3-methylphenol	0.321	0.297	7.5	105	0.00
37	2-Methylnaphthalene	0.698	0.651	6.7	106	0.00
38	1-Methylnaphthalene	0.656	0.617	5.9	106	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.674	0.647	4.0	106	0.00
41 P	Hexachlorocyclopentadiene	0.198	0.193	2.5	108	0.00
42 S	2,4,6-Tribromophenol	0.262	0.240	8.4	101	0.00
43 C	2,4,6-Trichlorophenol	0.426	0.405	4.9	105	0.00
44	2,4,5-Trichlorophenol	0.447	0.427	4.5	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022020\
 Data File : BF118966.D
 Acq On : 20 Feb 2020 18:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 21 02:10:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 20 17:40:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.358	1.259	7.3	105	0.00
46	1,1'-Biphenyl	1.616	1.538	4.8	104	0.00
47	2-Chloronaphthalene	1.303	1.255	3.7	106	0.00
48	2-Nitroaniline	0.363	0.339	6.6	103	0.00
49	Acenaphthylene	1.881	1.793	4.7	104	0.00
50	Dimethylphthalate	1.529	1.416	7.4	103	0.00
51	2,6-Dinitrotoluene	0.340	0.321	5.6	104	0.00
52 C	Acenaphthene	1.134	1.082	4.6	104	0.00
53	3-Nitroaniline	0.377	0.350	7.2	101	0.00
54 P	2,4-Dinitrophenol	0.140	0.129	7.9	100	0.00
55	Dibenzofuran	1.693	1.614	4.7	103	0.00
56 P	4-Nitrophenol	0.226	0.205	9.3	97	0.00
57	2,4-Dinitrotoluene	0.440	0.418	5.0	102	0.00
58	Fluorene	1.316	1.235	6.2	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.377	0.361	4.2	107	0.00
60	Diethylphthalate	1.441	1.328	7.8	101	0.00
61	4-Chlorophenyl-phenylether	0.696	0.659	5.3	103	0.00
62	4-Nitroaniline	0.385	0.354	8.1	100	0.00
63	Azobenzene	1.253	1.170	6.6	103	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.121	0.115	5.0	102	0.00
66 c	n-Nitrosodiphenylamine	0.655	0.637	2.7	103	0.00
67	4-Bromophenyl-phenylether	0.261	0.253	3.1	104	0.00
68	Hexachlorobenzene	0.301	0.292	3.0	103	0.00
69	Atrazine	0.229	0.212	7.4	99	0.00
70 C	Pentachlorophenol	0.121	0.114	5.8	101	0.00
71	Phenanthrene	1.091	1.044	4.3	101	0.00
72	Anthracene	1.090	1.034	5.1	99	0.00
73	Carbazole	0.971	0.928	4.4	101	0.00
74	Di-n-butylphthalate	1.176	1.100	6.5	99	0.00
75 C	Fluoranthene	1.158	1.082	6.6	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
77	Benzidine	0.813	0.780	4.1	92	0.00
78	Pyrene	1.606	1.538	4.2	97	0.00
79 S	Terphenyl-d14	1.162	1.109	4.6	98	0.00
80	Butylbenzylphthalate	0.676	0.626	7.4	92	0.00
81	Benzo(a)anthracene	1.363	1.281	6.0	92	0.00
82	3,3'-Dichlorobenzidine	0.529	0.495	6.4	88	0.00
83	Chrysene	1.358	1.276	6.0	93	0.00
84	Bis(2-ethylhexyl)phthalate	0.854	0.786	8.0	91	0.00
85 c	Di-n-octyl phthalate	1.361	1.254	7.9	86	0.00
86	Indeno(1,2,3-cd)pyrene	1.315	1.128	14.2	65	0.00
87 I	Perylene-d12	1.000	1.000	0.0	61	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022020\
Data File : BF118966.D
Acq On : 20 Feb 2020 18:39
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 21 02:10:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 20 17:40:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.318	1.257	4.6	88	0.00
89	Benzo(k)fluoranthene	1.222	1.230	-0.7	85	0.00
90 C	Benzo(a)pyrene	1.190	1.124	5.5	62	0.00
91	Dibenzo(a,h)anthracene	1.047	0.967	7.6	65	0.00
92	Benzo(q,h,i)perylene	1.034	0.957	7.4	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0