



Data Path : Z:\HPCHEM1\BNA F\DATA\BF022118\
 Data File : BF103126.D
 Acq On : 21 Feb 2018 10:37
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 22 02:59:37 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 16 14:34:49 2018
 Response via : Initial Calibration



TIC: BF103126.D

(76) Benzidine



12.739min (-0.006) 171331.53ng

response 850439

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.52
183.00	11.60	12.05
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022118\
Data File : BF103126.D
Acq On : 21 Feb 2018 10:37
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 22 02:59:37 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 16 14:34:49 2018
Response via : Initial Calibration



TIC: BF103126.D

(69) Pentachlorophenol (C)

11.210min (-0.000) 38792.94ng

response 143236

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	64.83
264.00	61.90	66.78
0.00	0.00	0.00