


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF022119\
 Data File : BF112644.D
 Acq On : 21 Feb 2019 11:22
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 28 10:38:18 2019



AutoFind: Scans 1615, 1616, 1617; Background Corrected with Scan 1610

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.1	30813	PASS
68	69	0.00	2	1.7	578	PASS
69	198	0.00	100	39.3	33565	PASS
70	69	0.00	2	0.9	307	PASS
127	198	10	80	53.5	45613	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	85325	PASS
199	198	5	9	7.0	5992	PASS
275	198	10	60	28.5	24293	PASS
365	198	1	100	4.3	3645	PASS
441	198	0.01	100	14.5	12338	PASS
442	198	50	100	86.7	73968	PASS
443	442	15	24	19.3	14251	PASS