

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022217\
 Data File : BF093088.D
 Acq On : 22 Feb 2017 21:10
 Operator : SJ/MA
 Sample : I1931-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P-2

Quant Time: Feb 23 00:31:06 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 17 17:31:55 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	154307	20.00	ng	-0.01
21) Naphthalene-d8	8.12	136	594047	20.00	ng	-0.02
38) Acenaphthene-d10	9.87	164	292136	20.00	ng	-0.02
63) Phenanthrene-d10	11.36	188	578597	20.00	ng	-0.02
75) Chrysene-d12	13.99	240	499034	20.00	ng	-0.03
86) Perylene-d12	15.41	264	394412	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	7399	0.82	ng	0.01
7) Phenol-d6	6.48	99	264768	22.88	ng	-0.01
23) Nitrobenzene-d5	7.41	82	912050	85.50	ng	-0.01
41) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
44) 2-Fluorobiphenyl	9.20	172	1380840	85.63	ng	-0.02
78) Terphenyl-d14	12.94	244	1474642	69.43	ng	-0.02
Target Compounds						
49) Dimethylphthalate	9.60	163	325034	16.52	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022217\
Data File : BF093088.D
Acq On : 22 Feb 2017 21:10
Operator : SJ/MA
Sample : I1931-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P-2

Quant Time: Feb 23 00:31:06 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 17:31:55 2017
Response via : Initial Calibration

