

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022221\
 Data File : BF123257.D
 Acq On : 22 Feb 2021 14:06
 Operator : JU/CG
 Sample : PB134758BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB134758BL

Quant Time: Feb 22 14:44:49 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	115617	20.00	ng	0.00
21) Naphthalene-d8	8.139	136	474489	20.00	ng	0.00
39) Acenaphthene-d10	9.898	164	247378	20.00	ng	0.00
64) Phenanthrene-d10	11.392	188	468024	20.00	ng	0.00
76) Chrysene-d12	14.039	240	396061	20.00	ng	0.00
86) Perylene-d12	15.515	264	372365	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	1021689	127.20	ng	0.02
7) Phenol-d6	6.492	99	1301258	119.25	ng	0.00
23) Nitrobenzene-d5	7.416	82	867665	83.71	ng	-0.01
42) 2,4,6-Tribromophenol	10.692	330	307239	121.97	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	1389538	80.69	ng	0.00
79) Terphenyl-d14	12.986	244	1543126	74.92	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022221\
Data File : BF123257.D
Acq On : 22 Feb 2021 14:06
Operator : JU/CG
Sample : PB134758BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB134758BL

Quant Time: Feb 22 14:44:49 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration

