

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022221\
 Data File : BF123266.D
 Acq On : 22 Feb 2021 19:15
 Operator : JU/CG
 Sample : M1452-28
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-7

Quant Time: Feb 23 03:51:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	110853	20.00	ng	0.00
21) Naphthalene-d8	8.139	136	630124	20.00	ng	# 0.00
39) Acenaphthene-d10	9.898	164	325901	20.00	ng	0.00
64) Phenanthrene-d10	11.386	188	610567	20.00	ng	# 0.00
76) Chrysene-d12	14.039	240	500432	20.00	ng	0.00
86) Perylene-d12	15.509	264	447515	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	593933	77.12	ng	0.01
7) Phenol-d6	6.486	99	730328	69.80	ng	0.00
23) Nitrobenzene-d5	7.416	82	659888	47.94	ng	-0.01
42) 2,4,6-Tribromophenol	10.692	330	301257	90.78	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	1133343	49.95	ng	0.00
79) Terphenyl-d14	12.986	244	1623655	62.39	ng	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022221\
Data File : BF123266.D
Acq On : 22 Feb 2021 19:15
Operator : JU/CG
Sample : M1452-28
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-7

Quant Time: Feb 23 03:51:05 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration

