

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022221\
 Data File : BF123271.D
 Acq On : 22 Feb 2021 21:58
 Operator : JU/CG
 Sample : M1425-13DL 25X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 41061DL

Quant Time: Feb 23 03:51:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	115651	20.00	ng	0.00
21) Naphthalene-d8	8.139	136	470379	20.00	ng	# 0.00
39) Acenaphthene-d10	9.898	164	241140	20.00	ng	0.00
64) Phenanthrene-d10	11.386	188	456961	20.00	ng	0.00
76) Chrysene-d12	14.039	240	330897	20.00	ng	0.00
86) Perylene-d12	15.509	264	259461	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	25213	3.14	ng	0.00
7) Phenol-d6	6.487	99	16953	1.55	ng	0.00
23) Nitrobenzene-d5	7.410	82	36221	3.52	ng	-0.02
42) 2,4,6-Tribromophenol	10.686	330	11657	4.75	ng	-0.01
45) 2-Fluorobiphenyl	9.216	172	59952	3.57	ng	0.00
79) Terphenyl-d14	12.974	244	52733	3.06	ng	-0.01
Target Compounds						
10) Phenol	6.498	94	37884	3.18	ng	Qvalue 96
20) 3+4-Methylphenols	7.263	107	238908	25.06	ng	# 60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022221\
Data File : BF123271.D
Acq On : 22 Feb 2021 21:58
Operator : JU/CG
Sample : M1425-13DL 25X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
41061DL

Quant Time: Feb 23 03:51:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration

