

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022315\
 Data File : BF077408.D
 Acq On : 23 Feb 2015 16:38
 Operator : TP/IZ
 Sample : G1345-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 UB10PX1

Quant Time: Feb 24 01:28:09 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 23 23:58:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.33	152	77586	20.00	ng	0.00
21) Naphthalene-d8	8.91	136	327149	20.00	ng	0.00
38) Acenaphthene-d10	11.09	164	168997	20.00	ng	0.00
63) Phenanthrene-d10	12.93	188	323619	20.00	ng	0.00
75) Chrysene-d12	16.24	240	348059	20.00	ng	-0.02
86) Perylene-d12	17.99	264	338730	20.00	ng	-0.11

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.63	112	538214	115.81	ng	0.01
7) Phenol-d6	6.89	99	662177	110.82	ng	0.00
23) Nitrobenzene-d5	8.02	82	404726	76.93	ng	-0.01
41) 2,4,6-Tribromophenol	12.08	330	194658	119.63	ng	0.01
44) 2-Fluorobiphenyl	10.26	172	935608	86.32	ng	0.00
78) Terphenyl-d14	14.93	244	1062453	71.36	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Phenol	6.91	94	20831	2.94	ng	91
49) Dimethylphthalate	10.76	163	64601	5.44	ng	99
70) Phenanthrene	12.97	178	102375	5.46	ng	99
74) Fluoranthene	14.44	202	84267	4.11	ng	97
77) Pyrene	14.73	202	73271	3.44	ng	98
80) Benzo(a)anthracene	16.23	228	51442	2.52	ng	# 84
82) Chrysene	16.27	228	82157	4.29	ng	# 89
87) Benzo(b)fluoranthene	17.53	252	65611m	3.33	ng	
89) Benzo(a)pyrene	17.92	252	39002	2.08	ng	# 55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022315\
Data File : BF077408.D
Acq On : 23 Feb 2015 16:38
Operator : TP/IZ
Sample : G1345-06
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
UB10PX1

Quant Time: Feb 24 01:28:09 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 23 23:58:25 2015
Response via : Initial Calibration

