

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022321\
 Data File : BF123273.D
 Acq On : 23 Feb 2021 10:20
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 23 11:00:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.802	0.797	0.6	93	-0.01
3	Pyridine	1.954	1.657	15.2	72	-0.01
4	n-Nitrosodimethylamine	0.903	0.787	12.8	76	-0.01
5 S	2-Fluorophenol	1.389	1.355	2.4	97	0.00
6	Aniline	2.230	2.172	2.6	96	0.00
7 S	Phenol-d6	1.888	1.849	2.1	97	0.00
8	2-Chlorophenol	1.493	1.485	0.5	97	0.00
9	Benzaldehyde	1.075	1.039	3.3	101	0.00
10 C	Phenol	2.057	1.979	3.8	96	0.00
11	bis(2-Chloroethyl)ether	1.476	1.443	2.2	99	0.00
12	1,3-Dichlorobenzene	1.553	1.546	0.5	97	0.00
13 C	1,4-Dichlorobenzene	1.582	1.586	-0.3	98	0.00
14	1,2-Dichlorobenzene	1.472	1.477	-0.3	98	0.00
15	Benzyl Alcohol	1.472	1.510	-2.6	99	0.00
16	2,2'-oxybis(1-Chloropropane	2.024	1.937	4.3	99	0.00
17	2-Methylphenol	1.256	1.227	2.3	97	0.00
18	Hexachloroethane	0.572	0.603	-5.4	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.123	1.173	-4.5	102	0.00
20	3+4-Methylphenols	1.649	1.549	6.1	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.519	0.515	0.8	100	0.00
23 S	Nitrobenzene-d5	0.437	0.440	-0.7	100	0.00
24	Nitrobenzene	0.461	0.461	0.0	100	0.00
25	Isophorone	0.817	0.818	-0.1	101	0.00
26 C	2-Nitrophenol	0.193	0.199	-3.1	99	0.00
27	2,4-Dimethylphenol	0.299	0.296	1.0	99	0.00
28	bis(2-Chloroethoxy)methane	0.428	0.428	0.0	101	0.00
29 C	2,4-Dichlorophenol	0.308	0.310	-0.6	100	0.00
30	1,2,4-Trichlorobenzene	0.335	0.336	-0.3	99	0.00
31	Naphthalene	1.096	1.081	1.4	98	0.00
32	Benzoic acid	0.220	0.202	8.2	86	0.00
33	4-Chloroaniline	0.446	0.448	-0.4	101	0.00
34 C	Hexachlorobutadiene	0.189	0.195	-3.2	102	0.00
35	Caprolactam	0.103	0.102	1.0	100	0.00
36 C	4-Chloro-3-methylphenol	0.367	0.369	-0.5	101	0.00
37	2-Methylnaphthalene	0.730	0.728	0.3	100	0.00
38	1-Methylnaphthalene	0.683	0.681	0.3	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.598	0.599	-0.2	101	0.00
41 P	Hexachlorocyclopentadiene	0.280	0.275	1.8	89	0.00
42 S	2,4,6-Tribromophenol	0.204	0.209	-2.5	101	0.00
43 C	2,4,6-Trichlorophenol	0.424	0.425	-0.2	99	0.00
44	2,4,5-Trichlorophenol	0.454	0.456	-0.4	101	0.00
45 S	2-Fluorobiphenyl	1.392	1.373	1.4	101	0.00
46	1,1'-Biphenyl	1.659	1.659	0.0	102	0.00
47	2-Chloronaphthalene	1.309	1.297	0.9	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022321\
 Data File : BF123273.D
 Acq On : 23 Feb 2021 10:20
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 23 11:00:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.469	0.481	-2.6	102	0.00
49	Acenaphthylene	1.986	1.966	1.0	101	0.00
50	Dimethylphthalate	1.496	1.497	-0.1	103	0.00
51	2,6-Dinitrotoluene	0.345	0.356	-3.2	102	0.00
52 C	Acenaphthene	1.369	1.369	0.0	101	0.00
53	3-Nitroaniline	0.383	0.377	1.6	98	0.00
54 P	2,4-Dinitrophenol	0.181	0.177	2.2	93	0.00
55	Dibenzofuran	1.854	1.839	0.8	102	0.00
56 P	4-Nitrophenol	0.305	0.300	1.6	93	0.00
57	2,4-Dinitrotoluene	0.464	0.475	-2.4	101	0.00
58	Fluorene	1.459	1.436	1.6	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.362	0.373	-3.0	100	0.00
60	Diethylphthalate	1.481	1.492	-0.7	102	0.00
61	4-Chlorophenyl-phenylether	0.694	0.693	0.1	102	0.00
62	4-Nitroaniline	0.385	0.384	0.3	99	0.00
63	Azobenzene	1.707	1.675	1.9	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.135	0.144	-6.7	98	0.00
66 c	n-Nitrosodiphenylamine	0.692	0.686	0.9	101	0.00
67	4-Bromophenyl-phenylether	0.218	0.220	-0.9	102	0.00
68	Hexachlorobenzene	0.230	0.232	-0.9	101	0.00
69	Atrazine	0.219	0.217	0.9	101	0.00
70 C	Pentachlorophenol	0.145	0.151	-4.1	98	0.00
71	Phenanthrene	1.183	1.154	2.5	100	0.00
72	Anthracene	1.180	1.173	0.6	101	0.00
73	Carbazole	1.111	1.094	1.5	101	0.00
74	Di-n-butylphthalate	1.243	1.231	1.0	102	0.00
75 C	Fluoranthene	1.254	1.232	1.8	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
77	Benzydine	0.646	0.622	3.7	101	0.00
78	Pyrene	1.561	1.588	-1.7	101	0.00
79 S	Terphenyl-d14	1.040	1.052	-1.2	100	0.00
80	Butylbenzylphthalate	0.660	0.672	-1.8	100	0.00
81	Benzo(a)anthracene	1.359	1.347	0.9	97	0.00
82	3,3'-Dichlorobenzidine	0.452	0.466	-3.1	101	0.00
83	Chrysene	1.335	1.333	0.1	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.909	0.900	1.0	100	0.00
85 c	Di-n-octyl phthalate	1.552	1.503	3.2	97	0.00
86 I	Perylene-d12	1.000	1.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	1.382	1.252	9.4	80	0.00
88	Benzo(b)fluoranthene	1.405	1.469	-4.6	93	0.00
89	Benzo(k)fluoranthene	1.274	1.365	-7.1	91	0.00
90 C	Benzo(a)pyrene	1.253	1.266	-1.0	89	0.00
91	Dibenzo(a,h)anthracene	1.186	1.074	9.4	81	0.00
92	Benzo(g,h,i)perylene	1.092	0.995	8.9	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022321\
Data File : BF123273.D
Acq On : 23 Feb 2021 10:20
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 23 11:00:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0