

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022324\
 Data File : BF137374.D
 Acq On : 23 Feb 2024 14:28
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF022324

Quant Time: Feb 24 04:26:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 24 04:23:49 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	40.112	-0.3	102	0.00
3	Pyridine	40.000	40.268	-0.7	104	0.00
4	n-Nitrosodimethylamine	40.000	40.019	-0.0	104	0.00
5 S	2-Fluorophenol	80.000	83.904	-4.9	103	0.00
6	Aniline	40.000	42.825	-7.1	104	0.00
7 S	Phenol-d6	80.000	80.823	-1.0	103	0.00
8	2-Chlorophenol	40.000	41.064	-2.7	105	0.00
9	Benzaldehyde	40.000	43.970	-9.9	104	0.00
10 C	Phenol	40.000	41.750	-4.4	105	0.00
11	bis(2-Chloroethyl)ether	40.000	40.023	-0.1	105	0.00
12	1,3-Dichlorobenzene	40.000	39.758	0.6	102	0.00
13 C	1,4-Dichlorobenzene	40.000	39.820	0.4	103	0.00
14	1,2-Dichlorobenzene	40.000	39.688	0.8	103	0.00
15	Benzyl Alcohol	40.000	44.499	-11.2	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.703	-1.8	104	0.00
17	2-Methylphenol	40.000	40.190	-0.5	102	0.00
18	Hexachloroethane	40.000	39.714	0.7	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.594	-4.0	107	0.00
20	3+4-Methylphenols	40.000	42.084	-5.2	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	40.439	-1.1	103	0.00
23 S	Nitrobenzene-d5	80.000	85.626	-7.0	107	0.00
24	Nitrobenzene	40.000	43.351	-8.4	106	0.00
25	Isophorone	40.000	40.036	-0.1	104	0.00
26 C	2-Nitrophenol	40.000	41.563	-3.9	111	0.00
27	2,4-Dimethylphenol	40.000	41.707	-4.3	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.449	-3.6	105	0.00
29 C	2,4-Dichlorophenol	40.000	42.024	-5.1	105	0.00
30	1,2,4-Trichlorobenzene	40.000	40.261	-0.7	103	0.00
31	Naphthalene	40.000	40.015	-0.0	104	0.00
32	Benzoic acid	40.000	41.851	-4.6	116	0.00
33	4-Chloroaniline	40.000	43.696	-9.2	106	0.00
34 C	Hexachlorobutadiene	40.000	39.390	1.5	103	0.00
35	Caprolactam	40.000	45.244	-13.1	113	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.949	-4.9	107	0.00
37	2-Methylnaphthalene	40.000	40.205	-0.5	104	0.00
38	1-Methylnaphthalene	40.000	40.317	-0.8	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.507	-1.3	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.862	-7.2	108	0.00
42 S	2,4,6-Tribromophenol	80.000	83.368	-4.2	108	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.037	-5.1	107	0.00
44	2,4,5-Trichlorophenol	40.000	41.648	-4.1	104	0.00
45 S	2-Fluorobiphenyl	80.000	77.575	3.0	104	0.00
46	1,1'-Biphenyl	40.000	40.039	-0.1	104	0.00
47	2-Chloronaphthalene	40.000	40.797	-2.0	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022324\
 Data File : BF137374.D
 Acq On : 23 Feb 2024 14:28
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF022324

Quant Time: Feb 24 04:26:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 24 04:23:49 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.544	-6.4	117	0.00
49	Acenaphthylene	40.000	40.400	-1.0	105	0.00
50	Dimethylphthalate	40.000	41.012	-2.5	107	0.00
51	2,6-Dinitrotoluene	40.000	46.056	-15.1	112	0.00
52 C	Acenaphthene	40.000	39.738	0.7	103	0.00
53	3-Nitroaniline	40.000	41.369	-3.4	109	0.00
54 P	2,4-Dinitrophenol	40.000	42.731	-6.8	119	0.00
55	Dibenzofuran	40.000	40.349	-0.9	106	0.00
56 P	4-Nitrophenol	40.000	42.238	-5.6	112	0.00
57	2,4-Dinitrotoluene	40.000	45.043	-12.6	116	0.00
58	Fluorene	40.000	39.504	1.2	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.685	-4.2	108	0.00
60	Diethylphthalate	40.000	41.599	-4.0	108	0.00
61	4-Chlorophenyl-phenylether	40.000	39.184	2.0	105	0.00
62	4-Nitroaniline	40.000	43.843	-9.6	117	0.00
63	Azobenzene	40.000	40.418	-1.0	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.591	-6.5	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.156	-0.4	106	0.00
67	4-Bromophenyl-phenylether	40.000	41.882	-4.7	108	0.00
68	Hexachlorobenzene	40.000	40.303	-0.8	106	0.00
69	Atrazine	40.000	44.062	-10.2	113	0.00
70 C	Pentachlorophenol	40.000	44.238	-10.6	108	0.00
71	Phenanthrene	40.000	39.656	0.9	107	0.00
72	Anthracene	40.000	40.158	-0.4	105	0.00
73	Carbazole	40.000	41.288	-3.2	107	0.00
74	Di-n-butylphthalate	40.000	46.745	-16.9	117	0.00
75 C	Fluoranthene	40.000	41.010	-2.5	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzidine	40.000	46.120	-15.3	131	0.00
78	Pyrene	40.000	41.279	-3.2	108	0.00
79 S	Terphenyl-d14	80.000	81.086	-1.4	109	0.00
80	Butylbenzylphthalate	40.000	48.688	-21.7	149	0.00
81	Benzo(a)anthracene	40.000	40.813	-2.0	112	0.00
82	3,3'-Dichlorobenzidine	40.000	42.858	-7.1	122	0.00
83	Chrysene	40.000	40.007	-0.0	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	47.819	-19.5	148	0.00
85 c	Di-n-octyl phthalate	40.000	45.252	-13.1	145	0.00
86 I	Perylene-d12	20.000	20.000	0.0	110	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.831	0.4	105	0.00
88	Benzo(b)fluoranthene	40.000	39.704	0.7	106	0.00
89	Benzo(k)fluoranthene	40.000	42.298	-5.7	116	0.00
90 C	Benzo(a)pyrene	40.000	41.767	-4.4	109	0.00
91	Dibenzo(a,h)anthracene	40.000	40.595	-1.5	106	0.00
92	Benzo(g,h,i)perylene	40.000	40.040	-0.1	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022324\
Data File : BF137374.D
Acq On : 23 Feb 2024 14:28
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF022324

Quant Time: Feb 24 04:26:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 24 04:23:49 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0