

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022420\
 Data File : BF119058.D
 Acq On : 25 Feb 2020 10:10
 Operator : CG/JU
 Sample : L1648-03
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PAX-2A

Quant Time: Feb 25 10:46:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 20 17:40:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	156715	20.00	ng	-0.01
21) Naphthalene-d8	8.03	136	609216	20.00	ng	-0.01
39) Acenaphthene-d10	9.79	164	337050	20.00	ng	0.00
64) Phenanthrene-d10	11.27	188	601370	20.00	ng	0.00
76) Chrysene-d12	13.90	240	414564	20.00	ng	0.00
87) Perylene-d12	15.33	264	396209	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	161896	17.26	ng	0.00
7) Phenol-d6	6.39	99	511123	44.82	ng	-0.01
23) Nitrobenzene-d5	7.31	82	569589	49.37	ng	-0.02
42) 2,4,6-Tribromophenol	10.57	330	19461	4.41	ng	0.00
45) 2-Fluorobiphenyl	9.11	172	1205463	52.69	ng	-0.01
79) Terphenyl-d14	12.85	244	1437021	59.68	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.50	163	61437	2.384	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022420\
Data File : BF119058.D
Acq On : 25 Feb 2020 10:10
Operator : CG/JU
Sample : L1648-03
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PAX-2A

Quant Time: Feb 25 10:46:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 20 17:40:17 2020
Response via : Initial Calibration

