




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022421\
Data File : BF123290.D
Acq On : 24 Feb 2021 12:37
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 24 13:33:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration



TIC: BF123290.D\data.ms

(70) Pentachlorophenol (C)




11.192min (-0.006) 22459.25 ng

response 173448

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	65.62
264.00	62.30	66.09
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022421\
Data File : BF123290.D
Acq On : 24 Feb 2021 12:37
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 24 13:33:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration



TIC: BF123290.D\data.ms

(77) Benzidine

12.733min (+ 0.000) 0.00 ng

response 1182472

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.82
183.00	12.10	12.77
0.00	0.00	0.00