

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF022421\  
 Data File : BF123291.D  
 Acq On : 24 Feb 2021 13:10  
 Operator : JU/CG  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: Feb 24 13:55:38 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF021921.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 19 16:08:53 2021  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | Amount | Calc.  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|--------|--------|-------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 20.000 | 20.000 | 0.0   | 94    | 0.00     |
| 2    | 1,4-Dioxane                 | 40.000 | 40.189 | -0.5  | 92    | -0.01    |
| 3    | Pyridine                    | 40.000 | 34.303 | 14.2  | 71    | -0.01    |
| 4    | n-Nitrosodimethylamine      | 40.000 | 35.296 | 11.8  | 76    | 0.00     |
| 5 S  | 2-Fluorophenol              | 80.000 | 78.638 | 1.7   | 95    | 0.00     |
| 6    | Aniline                     | 40.000 | 38.764 | 3.1   | 94    | 0.00     |
| 7 S  | Phenol-d6                   | 80.000 | 78.730 | 1.6   | 96    | 0.00     |
| 8    | 2-Chlorophenol              | 40.000 | 39.583 | 1.0   | 95    | 0.00     |
| 9    | Benzaldehyde                | 40.000 | 45.404 | -13.5 | 102   | 0.00     |
| 10 C | Phenol                      | 40.000 | 39.719 | 0.7   | 97    | 0.00     |
| 11   | bis(2-Chloroethyl)ether     | 40.000 | 39.231 | 1.9   | 97    | 0.00     |
| 12   | 1,3-Dichlorobenzene         | 40.000 | 39.672 | 0.8   | 95    | 0.00     |
| 13 C | 1,4-Dichlorobenzene         | 40.000 | 40.055 | -0.1  | 96    | 0.00     |
| 14   | 1,2-Dichlorobenzene         | 40.000 | 40.371 | -0.9  | 97    | 0.00     |
| 15   | Benzyl Alcohol              | 40.000 | 40.891 | -2.2  | 97    | 0.00     |
| 16   | 2,2'-oxybis(1-Chloropropane | 40.000 | 38.020 | 4.9   | 96    | 0.00     |
| 17   | 2-Methylphenol              | 40.000 | 39.601 | 1.0   | 96    | 0.00     |
| 18   | Hexachloroethane            | 40.000 | 42.070 | -5.2  | 98    | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine  | 40.000 | 41.517 | -3.8  | 99    | 0.00     |
| 20   | 3+4-Methylphenols           | 40.000 | 37.674 | 5.8   | 95    | 0.00     |
| 21 I | Naphthalene-d8              | 20.000 | 20.000 | 0.0   | 95    | 0.00     |
| 22   | Acetophenone                | 40.000 | 39.852 | 0.4   | 98    | 0.00     |
| 23 S | Nitrobenzene-d5             | 80.000 | 80.794 | -1.0  | 98    | 0.00     |
| 24   | Nitrobenzene                | 40.000 | 40.564 | -1.4  | 99    | 0.00     |
| 25   | Isophorone                  | 40.000 | 39.609 | 1.0   | 98    | 0.00     |
| 26 C | 2-Nitrophenol               | 40.000 | 40.822 | -2.1  | 96    | 0.00     |
| 27   | 2,4-Dimethylphenol          | 40.000 | 38.717 | 3.2   | 94    | 0.00     |
| 28   | bis(2-Chloroethoxy)methane  | 40.000 | 39.733 | 0.7   | 98    | 0.00     |
| 29 C | 2,4-Dichlorophenol          | 40.000 | 39.915 | 0.2   | 97    | 0.00     |
| 30   | 1,2,4-Trichlorobenzene      | 40.000 | 39.928 | 0.2   | 96    | 0.00     |
| 31   | Naphthalene                 | 40.000 | 39.511 | 1.2   | 96    | 0.00     |
| 32   | Benzoic acid                | 40.000 | 34.356 | 14.1  | 78    | 0.00     |
| 33   | 4-Chloroaniline             | 40.000 | 39.815 | 0.5   | 98    | 0.00     |
| 34 C | Hexachlorobutadiene         | 40.000 | 40.388 | -1.0  | 97    | 0.00     |
| 35   | Caprolactam                 | 40.000 | 39.650 | 0.9   | 98    | 0.00     |
| 36 C | 4-Chloro-3-methylphenol     | 40.000 | 40.954 | -2.4  | 100   | 0.00     |
| 37   | 2-Methylnaphthalene         | 40.000 | 38.977 | 2.6   | 95    | 0.00     |
| 38   | 1-Methylnaphthalene         | 40.000 | 39.476 | 1.3   | 97    | 0.00     |
| 39 I | Acenaphthene-d10            | 20.000 | 20.000 | 0.0   | 98    | 0.00     |
| 40   | 1,2,4,5-Tetrachlorobenzene  | 40.000 | 39.784 | 0.5   | 98    | 0.00     |
| 41 P | Hexachlorocyclopentadiene   | 40.000 | 38.139 | 4.7   | 85    | 0.00     |
| 42 S | 2,4,6-Tribromophenol        | 80.000 | 79.905 | 0.1   | 96    | 0.00     |
| 43 C | 2,4,6-Trichlorophenol       | 40.000 | 39.307 | 1.7   | 95    | 0.00     |
| 44   | 2,4,5-Trichlorophenol       | 40.000 | 38.910 | 2.7   | 95    | 0.00     |
| 45 S | 2-Fluorobiphenyl            | 80.000 | 77.527 | 3.1   | 97    | 0.00     |
| 46   | 1,1'-Biphenyl               | 40.000 | 39.153 | 2.1   | 97    | 0.00     |
| 47   | 2-Chloronaphthalene         | 40.000 | 39.065 | 2.3   | 97    | 0.00     |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF022421\  
 Data File : BF123291.D  
 Acq On : 24 Feb 2021 13:10  
 Operator : JU/CG  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: Feb 24 13:55:38 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF021921.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 19 16:08:53 2021  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | Amount | Calc.  | %Dev  | Area% | Dev(min) |
|------|----------------------------|--------|--------|-------|-------|----------|
| 48   | 2-Nitroaniline             | 40.000 | 41.054 | -2.6  | 99    | 0.00     |
| 49   | Acenaphthylene             | 40.000 | 38.859 | 2.9   | 97    | 0.00     |
| 50   | Dimethylphthalate          | 40.000 | 39.558 | 1.1   | 100   | 0.00     |
| 51   | 2,6-Dinitrotoluene         | 40.000 | 40.257 | -0.6  | 97    | 0.00     |
| 52 C | Acenaphthene               | 40.000 | 39.528 | 1.2   | 97    | 0.00     |
| 53   | 3-Nitroaniline             | 40.000 | 39.225 | 1.9   | 95    | 0.00     |
| 54 P | 2,4-Dinitrophenol          | 40.000 | 37.869 | 5.3   | 89    | 0.00     |
| 55   | Dibenzofuran               | 40.000 | 38.664 | 3.3   | 97    | 0.00     |
| 56 P | 4-Nitrophenol              | 40.000 | 37.438 | 6.4   | 87    | 0.00     |
| 57   | 2,4-Dinitrotoluene         | 40.000 | 40.425 | -1.1  | 97    | 0.00     |
| 58   | Fluorene                   | 40.000 | 38.850 | 2.9   | 97    | 0.00     |
| 59   | 2,3,4,6-Tetrachlorophenol  | 40.000 | 39.727 | 0.7   | 95    | 0.00     |
| 60   | Diethylphthalate           | 40.000 | 38.949 | 2.6   | 97    | 0.00     |
| 61   | 4-Chlorophenyl-phenylether | 40.000 | 39.412 | 1.5   | 98    | 0.00     |
| 62   | 4-Nitroaniline             | 40.000 | 39.464 | 1.3   | 96    | 0.00     |
| 63   | Azobenzene                 | 40.000 | 38.618 | 3.5   | 98    | 0.00     |
| 64 I | Phenanthrene-d10           | 20.000 | 20.000 | 0.0   | 96    | 0.00     |
| 65   | 4,6-Dinitro-2-methylphenol | 40.000 | 42.675 | -6.7  | 93    | 0.00     |
| 66 c | n-Nitrosodiphenylamine     | 40.000 | 39.481 | 1.3   | 96    | 0.00     |
| 67   | 4-Bromophenyl-phenylether  | 40.000 | 40.680 | -1.7  | 99    | 0.00     |
| 68   | Hexachlorobenzene          | 40.000 | 39.983 | 0.0   | 95    | 0.00     |
| 69   | Atrazine                   | 40.000 | 39.463 | 1.3   | 96    | 0.00     |
| 70 C | Pentachlorophenol          | 40.000 | 40.337 | -0.8  | 91    | 0.00     |
| 71   | Phenanthrene               | 40.000 | 39.054 | 2.4   | 96    | 0.00     |
| 72   | Anthracene                 | 40.000 | 39.372 | 1.6   | 96    | 0.00     |
| 73   | Carbazole                  | 40.000 | 39.256 | 1.9   | 96    | 0.00     |
| 74   | Di-n-butylphthalate        | 40.000 | 39.024 | 2.4   | 96    | 0.00     |
| 75 C | Fluoranthene               | 40.000 | 38.850 | 2.9   | 94    | 0.00     |
| 76 I | Chrysene-d12               | 20.000 | 20.000 | 0.0   | 89    | 0.00     |
| 77   | Benzydine                  | 40.000 | 35.120 | 12.2  | 85    | 0.00     |
| 78   | Pyrene                     | 40.000 | 40.921 | -2.3  | 94    | 0.00     |
| 79 S | Terphenyl-d14              | 80.000 | 83.021 | -3.8  | 95    | 0.00     |
| 80   | Butylbenzylphthalate       | 40.000 | 40.552 | -1.4  | 92    | 0.00     |
| 81   | Benzo(a)anthracene         | 40.000 | 39.981 | 0.0   | 91    | 0.00     |
| 82   | 3,3'-Dichlorobenzidine     | 40.000 | 39.478 | 1.3   | 89    | 0.00     |
| 83   | Chrysene                   | 40.000 | 40.031 | -0.1  | 91    | 0.00     |
| 84   | Bis(2-ethylhexyl)phthalate | 40.000 | 38.625 | 3.4   | 90    | 0.00     |
| 85 c | Di-n-octyl phthalate       | 40.000 | 36.864 | 7.8   | 85    | 0.00     |
| 86 I | Perylene-d12               | 20.000 | 20.000 | 0.0   | 78    | 0.00     |
| 87   | Indeno(1,2,3-cd)pyrene     | 40.000 | 35.565 | 11.1  | 70    | -0.01    |
| 88   | Benzo(b)fluoranthene       | 40.000 | 41.546 | -3.9  | 82    | 0.00     |
| 89   | Benzo(k)fluoranthene       | 40.000 | 44.678 | -11.7 | 84    | 0.00     |
| 90 C | Benzo(a)pyrene             | 40.000 | 40.493 | -1.2  | 79    | -0.01    |
| 91   | Dibenzo(a,h)anthracene     | 40.000 | 35.372 | 11.6  | 70    | -0.02    |
| 92   | Benzo(g,h,i)perylene       | 40.000 | 35.856 | 10.4  | 68    | -0.02    |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF022421\  
Data File : BF123291.D  
Acq On : 24 Feb 2021 13:10  
Operator : JU/CG  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_F  
LabSampleId :  
SSTDCCC040

Quant Time: Feb 24 13:55:38 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF021921.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 19 16:08:53 2021  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|----------|--------|-------|------|-------|----------|
|----------|--------|-------|------|-------|----------|

(#) = Out of Range                      SPCC's out = 0    CCC's out = 0