

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022515\
 Data File : BF077452.D
 Acq On : 26 Feb 2015 1:57
 Operator : IZ / TP
 Sample : G1350-01
 Misc : S
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC

Quant Time: Feb 26 05:00:08 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 26 04:51:14 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.30	152	75999	20.00	ng	0.01
21) Naphthalene-d8	8.88	136	280651	20.00	ng	0.01
38) Acenaphthene-d10	11.04	164	168414	20.00	ng	0.00
63) Phenanthrene-d10	12.88	188	324420	20.00	ng	0.00
75) Chrysene-d12	16.17	240	352964	20.00	ng	0.01
86) Perylene-d12	17.89	264	349296	20.00	ng	0.03


System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.58	112	340405	74.78	ng	0.01
7) Phenol-d6	6.85	99	506952	86.61	ng	0.01
23) Nitrobenzene-d5	8.00	82	313547	69.47	ng	0.02
41) 2,4,6-Tribromophenol	12.02	330	138894	85.65	ng	0.00
44) 2-Fluorobiphenyl	10.21	172	719830	66.64	ng	0.00
78) Terphenyl-d14	14.88	244	821967	54.44	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Aniline	6.93	93	26469	3.27	ng	# 1
9) Benzaldehyde	6.73	77	655803	184.55	ng	# 1
10) Phenol	6.86	94	23415	3.37	ng	81
15) Benzyl Alcohol	7.48	79	37849	8.51	ng	# 49
17) 2-Methylphenol	7.65	107	41126	9.80	ng	# 1
18) Hexachloroethane	7.92	117	401723	202.18	ng	# 1
19) n-Nitroso-di-n-propylamine	7.80	70	41924	11.28	ng	# 49
22) Acetophenone	7.79	105	1997467	302.57	ng	# 50
24) Nitrobenzene	7.97	77	422893	89.06	ng	# 16
26) 2-Nitrophenol	8.44	139	8274	4.25	ng	# 1
28) bis(2-Chloroethoxy)methane	8.57	93	33427	5.91	ng	# 95
29) 2,4-Dichlorophenol	8.73	162	438376	107.25	ng	# 27
31) Naphthalene	8.90	128	437722	31.19	ng	# 92
35) Caprolactam	9.37	113	5507	4.41	ng	# 1
37) 2-Methylnaphthalene	9.76	142	198153	19.88	ng	96
49) Dimethylphthalate	10.72	163	26433	2.23	ng	# 97
55) 4-Nitrophenol	11.17	139	16144	7.33	ng	# 68
56) 2,4-Dinitrotoluene	11.28	165	27800	8.67	ng	# 54
64) 4,6-Dinitro-2-methylphenol	11.81	198	1283	4.63	ng	# 1
65) n-Nitrosodiphenylamine	11.87	169	76856	7.14	ng	# 36
77) Pyrene	14.67	202	45540	2.11	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022515\  
Data File : BF077452.D  
Acq On    : 26 Feb 2015    1:57  
Operator  : IZ / TP  
Sample    : G1350-01  
Misc      : S  
ALS Vial  : 22    Sample Multiplier: 1
```

Quant Time: Feb 26 05:00:08 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 26 04:51:14 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 7.30 min Scan# 561

Delta R.T. 0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC


Tgt Ion:152 Resp: 75999

Ion Ratio Lower Upper

152 100

150 156.3 124.0 186.0




115 53.7 49.4 74.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 74.78 ng

RT: 5.58 min Scan# 411

Delta R.T. 0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57


Tgt Ion:112 Resp: 340405

Ion Ratio Lower Upper

112 100

64 60.2 48.5 72.7



63 43.2 25.0 37.4#




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#6

Aniline

Concen: 3.27 ng

RT: 6.93 min Scan# 529

Delta R.T. 0.06 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC


Tgt Ion: 93 Resp: 26469

Ion Ratio Lower Upper

93 100

66 98.5 29.1 43.7#




65 1001.9 14.6 22.0#



Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0



#7

Phenol-d6

Concen: 86.61 ng

RT: 6.85 min Scan# 522

Delta R.T. 0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57


Tgt Ion: 99 Resp: 506952

Ion Ratio Lower Upper

99 100

42 23.5 16.2 24.4



71 54.3 26.7 40.1#




Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0





#9

Benzaldehyde

Concen: 184.55 ng

RT: 6.73 min Scan# 511

Delta R.T. -0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC





Tgt Ion: 77 Resp: 655803

Ion Ratio Lower Upper

77 100

105 685.7 85.9 125.9#

106 71.8 82.9 122.9#



#10

Phenol

Concen: 3.37 ng

RT: 6.86 min Scan# 523

Delta R.T. 0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57




Tgt Ion: 94 Resp: 23415


Ion Ratio Lower Upper

94 100

65 41.0 5.7 45.7

66 39.8 13.6 53.6





#15

Benzyl Alcohol

Concen: 8.51 ng

RT: 7.48 min Scan# 577

Delta R.T. 0.02 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC





Tgt Ion: 79 Resp: 37849

Ion Ratio Lower Upper

79 100

108 14.5 58.3 87.5#

77 87.2 51.3 76.9#



#17

2-Methylphenol

Concen: 9.80 ng

RT: 7.65 min Scan# 592

Delta R.T. 0.05 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Tgt Ion: 107 Resp: 41126




Ion Ratio Lower Upper


107 100

108 3.7 91.2 136.8#

77 1483.9 33.6 50.4#

79 1183.2 32.1 48.1#









#18
Hexachloroethane
Concen: 202.18 ng
RT: 7.92 min Scan# 615
Delta R.T. -0.00 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57




Instrument :
BNA_F
ClientSampleId :
WC


Tgt Ion: 117 Resp: 401723
Ion Ratio Lower Upper
117 100
119 1075.0 77.8 116.8#
201 0.0 90.6 136.0#



#19
n-Nitroso-di-n-propylamine
Concen: 11.28 ng
RT: 7.80 min Scan# 605
Delta R.T. -0.00 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57

Tgt Ion: 70 Resp: 41924
Ion Ratio Lower Upper
70 100
42 105.9 48.3 72.5#
101 4.0 7.5 11.3#
130 0.8 16.7 25.1#





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.88 min Scan# 699

Delta R.T. 0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampled :

WC

Tgt Ion:136 Resp: 280651


Ion Ratio Lower Upper

136 100

137 11.3 8.6 13.0

54 9.0 6.2 9.4

68 6.6 5.0 7.6




Abundance Ion 136.00 (135.70 to 136.70): E


Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



Time-->



#22

Acetophenone

Concen: 302.57 ng

RT: 7.79 min Scan# 604

Delta R.T. -0.00 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Tgt Ion:105 Resp: 1997467


Ion Ratio Lower Upper

105 100

71 5.7 1.7 2.5#

51 4.5 30.7 46.1#

120 0.2 16.7 25.1#




Abundance Ion 105.00 (104.70 to 105.70): E


Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E



Time-->



#23

Nitrobenzene-d5

Concen: 69.47 ng

RT: 8.00 min Scan# 622

Delta R.T. 0.02 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC


Tgt Ion: 82 Resp: 313547

Ion Ratio Lower Upper




82 100

128 55.4 43.8 65.8

54 45.8 36.7 55.1



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): B
Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 89.06 ng

RT: 7.97 min Scan# 620

Delta R.T. -0.02 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57


Tgt Ion: 77 Resp: 422893

Ion Ratio Lower Upper



77 100


123 0.9 45.9 68.9#

65 58.6 9.5 14.3#



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 123.00 (122.70 to 123.70): B
Ion 65.00 (64.70 to 65.70): BF0





#26

2-Nitrophenol

Concen: 4.25 ng

RT: 8.44 min Scan# 661

Delta R.T. 0.05 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC





Tgt Ion: 139 Resp: 8274

Ion Ratio Lower Upper

139 100

109 224.2 22.6 34.0#

65 259.6 36.4 54.6#



#28

bis(2-Chloroethoxy)methane

Concen: 5.91 ng

RT: 8.57 min Scan# 672

Delta R.T. -0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57




Tgt Ion: 93 Resp: 33427


Ion Ratio Lower Upper

93 100

95 34.2 26.4 39.6

123 15.9 8.7 13.1#









#29
2,4-Dichlorophenol
Concen: 107.25 ng
RT: 8.73 min Scan# 686
Delta R.T. 0.03 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57

Instrument :
BNA_F
ClientSampleId :
WC

Tgt Ion:162 Resp: 438376
Ion Ratio Lower Upper
162 100
164 0.8 46.4 86.4#
98 1.2 6.9 46.9#




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): E






#31
Naphthalene
Concen: 31.19 ng
RT: 8.90 min Scan# 701
Delta R.T. 0.01 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57

Tgt Ion:128 Resp: 437722
Ion Ratio Lower Upper
128 100
129 16.4 9.4 14.0#
127 15.1 10.6 15.8



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E









#35
 Caprolactam
 Concen: 4.41 ng
 RT: 9.37 min Scan# 742
 Delta R.T. -0.01 min
 Lab File: BF077452.D
 Acq: 26 Feb 2015 1:57




Instrument :
 BNA_F
 ClientSampleID :
 WC


Tgt Ion:113 Resp: 5507
 Ion Ratio Lower Upper
 113 100
 55 795.3 154.6 194.6#
 56 422.9 114.0 154.0#



#37
 2-Methylnaphthalene
 Concen: 19.88 ng
 RT: 9.76 min Scan# 776
 Delta R.T. 0.01 min
 Lab File: BF077452.D
 Acq: 26 Feb 2015 1:57

Tgt Ion:142 Resp: 198153
 Ion Ratio Lower Upper
 142 100
 141 92.4 70.7 106.1
 115 31.6 24.2 36.4





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 11.04 min Scan# 888

Delta R.T. -0.00 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC


Tgt Ion:164 Resp: 168414

Ion Ratio Lower Upper

164 100

162 103.0 83.2 124.8




160 46.1 34.6 52.0



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#41

2,4,6-Tribromophenol

Concen: 85.65 ng

RT: 12.02 min Scan# 974

Delta R.T. -0.00 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57


Tgt Ion:330 Resp: 138894

Ion Ratio Lower Upper

330 100

332 94.2 77.4 116.2



141 37.7 28.8 43.2




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#44
2-Fluorobiphenyl
Concen: 66.64 ng
RT: 10.21 min Scan# 816
Delta R.T. -0.00 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57

Instrument :
BNA_F
ClientSampled :
WC

Tgt Ion:172 Resp: 719830
Ion Ratio Lower Upper
172 100
171 35.2 28.9 43.3
170 23.3 19.4 29.2




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#49
Dimethylphthalate
Concen: 2.23 ng
RT: 10.72 min Scan# 860
Delta R.T. -0.00 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57

Tgt Ion:163 Resp: 26433
Ion Ratio Lower Upper
163 100
194 5.7 2.6 4.0#
164 11.5 8.6 13.0



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#55
 4-Nitrophenol
 Concen: 7.33 ng
 RT: 11.17 min Scan# 900
 Delta R.T. -0.02 min
 Lab File: BF077452.D
 Acq: 26 Feb 2015 1:57




Instrument :
 BNA_F
 ClientSampleId :
 WC


Tgt Ion:139 Resp: 16144
 Ion Ratio Lower Upper
 139 100
 109 69.0 36.2 76.2
 65 25.9 40.3 80.3#



#56
 2,4-Dinitrotoluene
 Concen: 8.67 ng
 RT: 11.28 min Scan# 909
 Delta R.T. -0.00 min
 Lab File: BF077452.D
 Acq: 26 Feb 2015 1:57

Tgt Ion:165 Resp: 27800
 Ion Ratio Lower Upper
 165 100
 63 20.7 29.8 44.8#
 89 28.8 48.5 72.7#
 182 146.2 1.9 2.9#









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 12.88 min Scan# 1049
 Delta R.T. -0.00 min
 Lab File: BF077452.D
 Acq: 26 Feb 2015 1:57




Instrument :
 BNA_F
 ClientSampleId :
 WC


Tgt Ion:188 Resp: 324420
 Ion Ratio Lower Upper
 188 100
 94 11.3 7.4 11.2#
 80 12.1 8.0 12.0#



#64
 4,6-Dinitro-2-methylphenol
 Concen: 4.63 ng
 RT: 11.81 min Scan# 956
 Delta R.T. 0.03 min
 Lab File: BF077452.D
 Acq: 26 Feb 2015 1:57

Tgt Ion:198 Resp: 1283
 Ion Ratio Lower Upper
 198 100
 51 595.1 2.7 42.7#
 105 4619.7 10.0 50.0#





#65

n-Nitrosodiphenylamine

Concen: 7.14 ng

RT: 11.87 min Scan# 961

Delta R.T. -0.00 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57

Instrument :

BNA_F

ClientSampleId :

WC


Tgt Ion:169 Resp: 76856

Ion Ratio Lower Upper

169 100

168 6.9 53.0 79.6#




167 10.1 29.5 44.3#



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#75

Chrysene-d12

Concen: 20.00 ng

RT: 16.17 min Scan# 1337

Delta R.T. 0.01 min

Lab File: BF077452.D

Acq: 26 Feb 2015 1:57


Tgt Ion:240 Resp: 352964

Ion Ratio Lower Upper

240 100

120 11.0 8.8 13.2



236 26.4 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#77
Pyrene
Concen: 2.11 ng
RT: 14.67 min Scan# 1206
Delta R.T. 0.01 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57




Instrument :
BNA_F
ClientSampleId :
WC


Tgt Ion: 202 Resp: 45540
Ion Ratio Lower Upper
202 100
200 23.0 17.1 25.7
203 31.3 14.2 21.2#



#78
Terphenyl-d14
Concen: 54.44 ng
RT: 14.88 min Scan# 1224
Delta R.T. -0.00 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57

Tgt Ion: 244 Resp: 821967
Ion Ratio Lower Upper
244 100
212 6.9 5.2 7.8
122 11.3 7.8 11.8






#86
Perylene-d12
Concen: 20.00 ng
RT: 17.89 min Scan# 1488
Delta R.T. 0.03 min
Lab File: BF077452.D
Acq: 26 Feb 2015 1:57

Instrument :
BNA_F
ClientSampleId :
WC

Tgt Ion: 264 Resp: 349296
Ion Ratio Lower Upper
264 100
260 23.7 19.3 28.9
265 22.0 16.8 25.2



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E

