

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022519\
 Data File : BF112690.D
 Acq On : 25 Feb 2019 14:14
 Operator : JU/SJ
 Sample : K1595-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-E-WALL-GRADE-BEAM

Manual Integrations
 APPROVED

Sohil
 2/26/2019 2:39:09 PM

Quant Time: Feb 25 15:43:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 25 14:05:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	49634	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	188485	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	87159	20.00	ng	0.00
64) Phenanthrene-d10	11.34	188	136698	20.00	ng	0.00
76) Chrysene-d12	13.98	240	116016	20.00	ng	0.00
87) Perylene-d12	15.46	264	105230	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	264023	112.99	ng	0.00
7) Phenol-d6	6.48	99	338233	104.17	ng	-0.01
23) Nitrobenzene-d5	7.38	82	211601	64.69	ng	-0.01
42) 2,4,6-Tribromophenol	10.65	330	79975	78.83	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	400620	65.01	ng	0.00
79) Terphenyl-d14	12.91	244	281512	45.70	ng	-0.01
Target Compounds						
10) Phenol	6.50	94	10106	2.837	ng	Qvalue # 57
50) Dimethylphthalate	9.56	163	30312	4.476	ng	99
71) Phenanthrene	11.36	178	131294	18.475	ng	97
72) Anthracene	11.42	178	27744m	3.919	ng	
75) Fluoranthene	12.56	202	144534	18.962	ng	94
78) Pyrene	12.79	202	127705	15.899	ng	96
80) Butylbenzylphthalate	13.37	149	23758	6.113	ng	89
81) Benzo(a)anthracene	13.97	228	65499	9.604	ng	96
83) Chrysene	14.01	228	61267	9.411	ng	96
86) Indeno(1,2,3-cd)pyrene	16.96	276	32313	6.264	ng	# 90
88) Benzo(b)fluoranthene	15.03	252	82499m	13.109	ng	
89) Benzo(k)fluoranthene	15.05	252	25381m	4.211	ng	
90) Benzo(a)pyrene	15.40	252	54605	9.643	ng	96
91) Dibenzo(a,h)anthracene	16.95	278	9719	2.130	ng	# 89
92) Benzo(a,h,i)perylene	17.42	276	29664	6.890	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022519\
Data File : BF112690.D
Acq On : 25 Feb 2019 14:14
Operator : JU/SJ
Sample : K1595-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-E-WALL-GRADE-BEAM

Manual Integrations
APPROVED

Sohil
2/26/2019 2:39:09 PM

Quant Time: Feb 25 15:43:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 25 14:05:25 2019
Response via : Initial Calibration

