

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022519\
 Data File : BF112703.D
 Acq On : 25 Feb 2019 20:08
 Operator : JU/SJ
 Sample : K1595-06RE
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-E-WALL-GRADE-BEAMRE

Quant Time: Feb 26 00:30:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	51214	20.00	ng	-0.03
21) Naphthalene-d8	8.10	136	198351	20.00	ng	-0.03
39) Acenaphthene-d10	9.85	164	92077	20.00	ng	-0.04
64) Phenanthrene-d10	11.34	188	157295	20.00	ng	-0.03
76) Chrysene-d12	13.99	240	124373	20.00	ng	-0.02
87) Perylene-d12	15.46	264	94767	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	379111	157.23	ng	-0.02
7) Phenol-d6	6.48	99	419807	125.30	ng	-0.01
23) Nitrobenzene-d5	7.38	82	335509	97.46	ng	-0.03
42) 2,4,6-Tribromophenol	10.65	330	140505	131.10	ng	-0.04
45) 2-Fluorobiphenyl	9.17	172	655951	100.76	ng	-0.04
79) Terphenyl-d14	12.91	244	617343	93.49	ng	-0.04


Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022519\
Data File : BF112703.D
Acq On : 25 Feb 2019 20:08
Operator : JU/SJ
Sample : K1595-06RE
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-E-WALL-GRADE-BEAMRE

Quant Time: Feb 26 00:30:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.82 min Scan# 804
Delta R.T. -0.03 min
Lab File: BF112703.D
Acq: 25 Feb 2019 20:08

Instrument :
BNA_F
ClientSampleId :
TP-E-WALL-GRADE-BEAMRE

Tot Ion:152 Resp: 51214
Ion Ratio Lower Upper
152 100
150 157.8 127.4 191.0
115 58.6 45.5 68.3




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 157.234 ng
RT: 5.46 min Scan# 574
Delta R.T. -0.02 min
Lab File: BF112703.D
Acq: 25 Feb 2019 20:08

Tgt Ion:112 Resp: 379111
Ion Ratio Lower Upper
112 100
64 57.5 45.7 68.5
63 31.1 23.6 35.4



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 125.301 ng

RT: 6.48 min Scan# 747

Delta R.T. -0.01 min

Lab File: BF112703.D

Acq: 25 Feb 2019 20:08

Instrument :

BNA_F

ClientSampleId :

TP-E-WALL-GRADE-BEAMRE


Tot Ion: 99 Resp: 419807

Ion Ratio Lower Upper




99 100

42 20.3 15.1 22.7

71 38.3 26.9 40.3



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.10 min Scan# 1022

Delta R.T. -0.03 min

Lab File: BF112703.D

Acq: 25 Feb 2019 20:08

Tgt Ion: 136 Resp: 198351


Ion Ratio Lower Upper

136 100



137 10.8 9.1 13.7


54 5.9 5.9 8.9#

68 4.2 5.1 7.7#



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1









#23
Nitrobenzene-d5
Concen: 97.463 ng
RT: 7.38 min Scan# 900
Delta R.T. -0.03 min
Lab File: BF112703.D
Acq: 25 Feb 2019 20:08

Instrument :
BNA_F
ClientSampleId :
TP-E-WALL-GRADE-BEAMRE

Tot Ion: 82 Resp: 335509
Ion Ratio Lower Upper
82 100
128 43.3 37.8 56.8
54 50.3 39.7 59.5




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.85 min Scan# 1320
Delta R.T. -0.04 min
Lab File: BF112703.D
Acq: 25 Feb 2019 20:08

Tgt Ion:164 Resp: 92077
Ion Ratio Lower Upper
164 100
162 101.2 82.2 123.4
160 47.2 36.2 54.4



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E









#42
 2,4,6-Tribromophenol
 Concen: 131.099 ng
 RT: 10.65 min Scan# 1455
 Delta R.T. -0.04 min
 Lab File: BF112703.D
 Acq: 25 Feb 2019 20:08

Instrument :
 BNA_F
 ClientSampleId :
 TP-E-WALL-GRADE-BEAMRE

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.0	76.6	114.8
141	35.2	24.3	36.5




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E





#45
 2-Fluorobiphenyl
 Concen: 100.757 ng
 RT: 9.17 min Scan# 1204
 Delta R.T. -0.04 min
 Lab File: BF112703.D
 Acq: 25 Feb 2019 20:08

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.9	30.5	45.7
170	23.5	20.2	30.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.34 min Scan# 1573

Delta R.T. -0.03 min

Lab File: BF112703.D

Acq: 25 Feb 2019 20:08

Instrument :

BNA_F

ClientSampleId :

TP-E-WALL-GRADE-BEAMRE





Tot Ion:188 Resp: 157295

Ion Ratio Lower Upper

188 100

94 7.4 8.2 12.2#

80 8.8 8.9 13.3#



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.99 min Scan# 2023

Delta R.T. -0.02 min

Lab File: BF112703.D

Acq: 25 Feb 2019 20:08




Tgt Ion:240 Resp: 124373


Ion Ratio Lower Upper

240 100

120 9.5 9.0 13.6

236 24.7 20.7 31.1









#79
 Terphenyl-d14
 Concen: 93.487 ng
 RT: 12.91 min Scan# 1840
 Delta R.T. -0.04 min
 Lab File: BF112703.D
 Acq: 25 Feb 2019 20:08

Instrument :
 BNA_F
 ClientSampleId :
 TP-E-WALL-GRADE-BEAMRE

Tot Ion:244 Resp: 617343
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.9 8.9
 122 12.0 9.8 14.6



#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.46 min Scan# 2273
 Delta R.T. -0.02 min
 Lab File: BF112703.D
 Acq: 25 Feb 2019 20:08

Tgt Ion:264 Resp: 94767
 Ion Ratio Lower Upper
 264 100
 260 25.8 18.2 27.2
 265 21.6 17.0 25.4

