

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022525\
 Data File : BF141760.D
 Acq On : 25 Feb 2025 13:02
 Operator : RC/JU
 Sample : PB166824TB
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB166824TB

Quant Time: Feb 25 13:26:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 16:58:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	83192	20.000	ng	-0.01
21) Naphthalene-d8	8.063	136	337595	20.000	ng	-0.01
39) Acenaphthene-d10	9.816	164	183626	20.000	ng	-0.02
64) Phenanthrene-d10	11.304	188	308241	20.000	ng	-0.01
76) Chrysene-d12	13.939	240	184115	20.000	ng	-0.02
86) Perylene-d12	15.386	264	164298	20.000	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.410	112	780174	147.023	ng	0.00
7) Phenol-d6	6.428	99	942678	140.552	ng	-0.01
23) Nitrobenzene-d5	7.345	82	630232	97.370	ng	-0.02
42) 2,4,6-Tribromophenol	10.610	330	246569	133.670	ng	-0.02
45) 2-Fluorobiphenyl	9.139	172	1206046	101.189	ng	-0.01
79) Terphenyl-d14	12.886	244	1254306	115.009	ng	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022525\
Data File : BF141760.D
Acq On : 25 Feb 2025 13:02
Operator : RC/JU
Sample : PB166824TB
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB166824TB

Quant Time: Feb 25 13:26:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 06 16:58:59 2025
Response via : Initial Calibration

