

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
 Data File : BF077477.D
 Acq On : 26 Feb 2015 20:57
 Operator : TP/IZ
 Sample : PB81894BL
 Misc : W/DOD
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB81894BL

Quant Time: Feb 27 03:17:02 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 26 17:14:46 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.26	152	81408	20.00	ng	0.00
21) Naphthalene-d8	8.84	136	336488	20.00	ng	0.00
38) Acenaphthene-d10	11.01	164	174965	20.00	ng	-0.01
63) Phenanthrene-d10	12.87	188	356804	20.00	ng	0.00
75) Chrysene-d12	16.16	240	395307	20.00	ng	-0.02
86) Perylene-d12	17.89	264	394611	20.00	ng	-0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	443986	91.05	ng	0.00
7) Phenol-d6	6.82	99	513206	81.85	ng	-0.01
23) Nitrobenzene-d5	7.96	82	315900	58.38	ng	0.00
41) 2,4,6-Tribromophenol	12.00	330	158249	93.93	ng	-0.01
44) 2-Fluorobiphenyl	10.19	172	743132	66.23	ng	0.00
78) Terphenyl-d14	14.85	244	910933	53.87	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
Data File : BF077477.D
Acq On : 26 Feb 2015 20:57
Operator : TP/IZ
Sample : PB81894BL
Misc : W/DOD
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB81894BL

Quant Time: Feb 27 03:17:02 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 26 17:14:46 2015
Response via : Initial Calibration

