

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
 Data File : BF077486.D
 Acq On : 27 Feb 2015 3:07
 Operator : TP/IZ
 Sample : G1385-07 5X
 Misc : S
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-9-D4

Manual Integrations
 APPROVED

apatel
 2/27/2015 4:21:58 PM

Quant Time: Feb 27 07:26:46 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 27 07:02:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.27	152	73055	20.00	ng	0.00
21) Naphthalene-d8	8.84	136	295618	20.00	ng	0.00
38) Acenaphthene-d10	11.02	164	155480	20.00	ng	0.00
63) Phenanthrene-d10	12.86	188	300701	20.00	ng	0.00
75) Chrysene-d12	16.15	240	338949	20.00	ng	0.00
86) Perylene-d12	17.88	264	355594	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.55	112	56180	12.84	ng	0.00
7) Phenol-d6	6.82	99	79629	14.15	ng	0.00
23) Nitrobenzene-d5	7.95	82	44174	9.29	ng	0.00
41) 2,4,6-Tribromophenol	12.00	330	17035	11.38	ng	0.00
44) 2-Fluorobiphenyl	10.19	172	107705	10.80	ng	0.00
78) Terphenyl-d14	14.86	244	128949	8.89	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.87	128	139694	9.45	ng	99
37) 2-Methylnaphthalene	9.73	142	36221	3.45	ng	97
51) Acenaphthene	11.06	154	101441	11.62	ng	98
54) Dibenzofuran	11.27	168	103667	7.69	ng	94
57) Fluorene	11.70	166	120077	11.15	ng	98
70) Phenanthrene	12.89	178	1246933	71.55	ng	98
71) Anthracene	12.95	178	337365	19.51	ng	99
72) Carbazole	13.16	167	177967	10.82	ng	99
74) Fluoranthene	14.37	202	1559672	81.92	ng	98
77) Pyrene	14.65	202	1268849	61.20	ng	96
80) Benzo(a)anthracene	16.14	228	720078	36.25	ng	97
82) Chrysene	16.18	228	602567	32.30	ng	97
83) Bis(2-ethylhexyl)phthalate	16.20	149	42371	3.10	ng	# 94
85) Indeno(1,2,3-cd)pyrene	19.33	276	336738	15.83	ng	96
87) Benzo(b)fluoranthene	17.45	252	1020134	49.33	ng	# 96
88) Benzo(k)fluoranthene	17.47	252	159083m	7.85	ng	98
89) Benzo(a)pyrene	17.82	252	545192	27.68	ng	98
90) Dibenzo(a,h)anthracene	19.35	278	76283	4.06	ng	# 91
91) Benzo(g,h,i)perylene	19.76	276	325957	17.19	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
Data File : BF077486.D
Acq On : 27 Feb 2015 3:07
Operator : TP/IZ
Sample : G1385-07 5X
Misc : S
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-9-D4

Manual Integrations
APPROVED

apatel
2/27/2015 4:21:58 PM

Quant Time: Feb 27 07:26:46 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 27 07:02:57 2015
Response via : Initial Calibration

