

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
 Data File : BF077491.D
 Acq On : 27 Feb 2015 5:29
 Operator : TP/IZ
 Sample : G1385-13
 Misc : S
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-13-D9

Manual Integrations
 APPROVED

apatel
 2/27/2015 4:22:05 PM

Quant Time: Feb 27 11:33:35 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 27 07:02:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.27	152	76301	20.00	ng	0.00
21) Naphthalene-d8	8.84	136	310674	20.00	ng	0.00
38) Acenaphthene-d10	11.02	164	161565	20.00	ng	0.00
63) Phenanthrene-d10	12.86	188	349019	20.00	ng	0.00
75) Chrysene-d12	16.16	240	367274	20.00	ng	0.00
86) Perylene-d12	17.90	264	335297	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	318769	69.75	ng	0.01
7) Phenol-d6	6.83	99	423018	71.99	ng	0.00
23) Nitrobenzene-d5	7.96	82	247125	49.46	ng	0.00
41) 2,4,6-Tribromophenol	12.00	330	111727	71.82	ng	0.00
44) 2-Fluorobiphenyl	10.19	172	471166	45.47	ng	0.00
78) Terphenyl-d14	14.86	244	610900	38.89	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	10.70	163	23447	2.06	ng	# 99
70) Phenanthrene	12.89	178	73625	3.64	ng	98
74) Fluoranthene	14.37	202	155758	7.05	ng	97
77) Pyrene	14.65	202	137433	6.12	ng	99
80) Benzo(a)anthracene	16.14	228	49766	2.31	ng	99
82) Chrysene	16.18	228	60991	3.02	ng	98
87) Benzo(b)fluoranthene	17.45	252	73415m	3.77	ng	
89) Benzo(a)pyrene	17.83	252	41577	2.24	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
Data File : BF077491.D
Acq On : 27 Feb 2015 5:29
Operator : TP/IZ
Sample : G1385-13
Misc : S
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-13-D9

Manual Integrations
APPROVED

apatel
2/27/2015 4:22:05 PM

Quant Time: Feb 27 11:33:35 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 27 07:02:57 2015
Response via : Initial Calibration

