

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
 Data File : BF077492.D
 Acq On : 27 Feb 2015 5:58
 Operator : TP/IZ
 Sample : G1385-14
 Misc : S
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-14-D910

Quant Time: Feb 27 10:58:53 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 27 07:02:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.27	152	72955	20.00	ng	0.00
21) Naphthalene-d8	8.85	136	285043	20.00	ng	0.00
38) Acenaphthene-d10	11.02	164	152979	20.00	ng	0.00
63) Phenanthrene-d10	12.87	188	319273	20.00	ng	0.00
75) Chrysene-d12	16.18	240	383355	20.00	ng	0.03
86) Perylene-d12	18.00	264	344192	20.00	ng	0.12
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	352132	80.58	ng	0.01
7) Phenol-d6	6.83	99	464338	82.64	ng	0.00
23) Nitrobenzene-d5	7.96	82	292090	63.72	ng	0.00
41) 2,4,6-Tribromophenol	12.00	330	112874	76.63	ng	0.00
44) 2-Fluorobiphenyl	10.19	172	544595	55.51	ng	0.00
78) Terphenyl-d14	14.86	244	745431	45.46	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.87	128	415722	29.16	ng	99
37) 2-Methylnaphthalene	9.73	142	212482	20.99	ng	99
49) Dimethylphthalate	10.70	163	25524	2.37	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF022615\
Data File : BF077492.D
Acq On : 27 Feb 2015 5:58
Operator : TP/IZ
Sample : G1385-14
Misc : S
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-14-D910

Quant Time: Feb 27 10:58:53 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 27 07:02:57 2015
Response via : Initial Calibration

