

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022618\  
 Data File : BF103229.D  
 Acq On : 26 Feb 2018 20:38  
 Operator : SJ/JU  
 Sample : J1691-33 2X  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 B9

Quant Time: Feb 27 00:53:35 2018  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 23 02:43:46 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.87  | 152  | 161092   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.16  | 136  | 648579   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10        | 9.91  | 164  | 259585   | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10        | 11.40 | 188  | 386782   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12            | 14.05 | 240  | 294137   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12            | 15.55 | 264  | 201541   | 20.00 | ng    | 0.01     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.49  | 112  | 318897   | 29.94 | ng    | 0.00     |
| 7) Phenol-d6                | 6.50  | 99   | 376305   | 28.87 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.43  | 82   | 211968   | 18.66 | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 10.70 | 330  | 52186    | 23.75 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.23  | 172  | 378836   | 19.67 | ng    | 0.00     |
| 78) Terphenyl-d14           | 12.99 | 244  | 269539   | 22.03 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022618\  
Data File : BF103229.D  
Acq On : 26 Feb 2018 20:38  
Operator : SJ/JU  
Sample : J1691-33 2X  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
B9

Quant Time: Feb 27 00:53:35 2018  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 23 02:43:46 2018  
Response via : Initial Calibration

