

Data Path : Z:\HPCHEM1\BNA F\Data\BF022618\
 Data File : BF103205.D
 Acq On : 26 Feb 2018 9:09
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 26 10:40:46 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 02:43:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.698	0.714	-2.3	107	0.00
3	Pyridine	1.865	1.875	-0.5	104	0.02
4	n-Nitrosodimethylamine	0.752	0.776	-3.2	108	0.02
5 S	2-Fluorophenol	1.322	1.307	1.1	105	0.00
6	Aniline	2.335	2.323	0.5	106	0.00
7 S	Phenol-d6	1.618	1.570	3.0	105	0.00
8	2-Chlorophenol	1.374	1.360	1.0	105	0.00
9	Benzaldehyde	1.014	0.957	5.6	103	0.00
10 C	Phenol	1.878	1.855	1.2	106	0.00
11	bis(2-Chloroethyl)ether	1.426	1.418	0.6	106	0.00
12	1,3-Dichlorobenzene	1.570	1.545	1.6	106	0.00
13 C	1,4-Dichlorobenzene	1.574	1.540	2.2	105	0.00
14	1,2-Dichlorobenzene	1.451	1.411	2.8	105	0.00
15	Benzyl Alcohol	1.219	1.192	2.2	105	0.00
16	2,2'-oxybis(1-Chloropropane	2.448	2.404	1.8	103	0.00
17	2-Methylphenol	1.248	1.236	1.0	106	0.00
18	Hexachloroethane	0.573	0.569	0.7	106	0.00
19 P	n-Nitroso-di-n-propylamine	1.007	0.945	6.2	105	0.00
20	3+4-Methylphenols	1.522	1.405	7.7	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.467	0.444	4.9	105	0.00
23 S	Nitrobenzene-d5	0.350	0.343	2.0	106	0.00
24	Nitrobenzene	0.386	0.381	1.3	105	0.00
25	Isophorone	0.690	0.663	3.9	105	0.00
26 C	2-Nitrophenol	0.182	0.184	-1.1	103	0.00
27	2,4-Dimethylphenol	0.277	0.270	2.5	104	0.00
28	bis(2-Chloroethoxy)methane	0.406	0.395	2.7	105	0.00
29 C	2,4-Dichlorophenol	0.266	0.260	2.3	105	0.00
30	1,2,4-Trichlorobenzene	0.283	0.276	2.5	104	0.00
31	Naphthalene	0.979	0.936	4.4	104	0.00
32	Benzoic acid	0.183	0.235	-28.4#	129	0.00
33	4-Chloroaniline	0.423	0.410	3.1	103	0.00
34 C	Hexachlorobutadiene	0.147	0.143	2.7	106	0.00
35	Caprolactam	0.093	0.093	0.0	103	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.292	2.7	104	0.00
37	2-Methylnaphthalene	0.633	0.607	4.1	105	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
39	1,2,4,5-Tetrachlorobenzene	0.547	0.533	2.6	104	0.00
40 P	Hexachlorocyclopentadiene	0.152	0.142	6.6	95	0.00
41 S	2,4,6-Tribromophenol	0.169	0.161	4.7	98	0.00
42 C	2,4,6-Trichlorophenol	0.368	0.362	1.6	103	0.00
43	2,4,5-Trichlorophenol	0.423	0.410	3.1	101	0.00
44 S	2-Fluorobiphenyl	1.177	1.139	3.2	104	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF022618\
 Data File : BF103205.D
 Acq On : 26 Feb 2018 9:09
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 26 10:40:46 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 02:43:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.676	1.581	5.7	104	0.00
46	2-Chloronaphthalene	1.326	1.273	4.0	104	0.00
47	2-Nitroaniline	0.491	0.487	0.8	103	0.00
48	Acenaphthylene	2.040	1.909	6.4	101	0.00
49	Dimethylphthalate	1.604	1.509	5.9	102	0.00
50	2,6-Dinitrotoluene	0.337	0.325	3.6	101	0.00
51 C	Acenaphthene	1.190	1.132	4.9	104	0.00
52	3-Nitroaniline	0.427	0.420	1.6	104	0.00
53 P	2,4-Dinitrophenol	0.094	0.090	4.3	96	0.00
54	Dibenzofuran	1.633	1.579	3.3	101	0.00
55 P	4-Nitrophenol	0.265	0.236	10.9	92	0.00
56	2,4-Dinitrotoluene	0.426	0.423	0.7	101	0.00
57	Fluorene	1.391	1.252	10.0	103	0.00
58	2,3,4,6-Tetrachlorophenol	0.284	0.278	2.1	102	0.00
59	Diethylphthalate	1.535	1.440	6.2	102	0.00
60	4-Chlorophenyl-phenylether	0.590	0.562	4.7	103	0.00
61	4-Nitroaniline	0.427	0.412	3.5	100	0.00
62	Azobenzene	1.562	1.502	3.8	103	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
64	4,6-Dinitro-2-methylphenol	0.112	0.110	1.8	97	0.00
65 c	n-Nitrosodiphenylamine	0.696	0.663	4.7	102	0.00
66	4-Bromophenyl-phenylether	0.208	0.197	5.3	99	0.00
67	Hexachlorobenzene	0.226	0.215	4.9	102	0.00
68	Atrazine	0.209	0.194	7.2	100	0.00
69 C	Pentachlorophenol	0.109	0.104	4.6	98	0.00
70	Phenanthrene	1.101	1.021	7.3	101	0.00
71	Anthracene	1.129	1.064	5.8	102	0.00
72	Carbazole	1.124	1.046	6.9	100	0.00
73	Di-n-butylphthalate	1.406	1.321	6.0	101	0.00
74 C	Fluoranthene	1.106	1.017	8.0	99	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
76	Benzidine	0.748	0.629	15.9	87	0.00
77	Pyrene	1.559	1.549	0.6	99	0.00
78 S	Terphenyl-d14	0.832	0.787	5.4	99	0.00
79	Butylbenzylphthalate	0.824	0.819	0.6	97	0.00
80	Benzo(a)anthracene	1.298	1.265	2.5	96	0.00
81	3,3'-Dichlorobenzidine	0.463	0.445	3.9	93	0.00
82	Chrysene	1.260	1.203	4.5	95	0.00
83	Bis(2-ethylhexyl)phthalate	1.112	1.078	3.1	95	0.00
84 c	Di-n-octyl phthalate	1.796	1.674	6.8	92	0.00
85	Indeno(1,2,3-cd)pyrene	0.998	1.023	-2.5	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.00
87	Benzo(b)fluoranthene	1.315	1.220	7.2	91	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF022618\
Data File : BF103205.D
Acq On : 26 Feb 2018 9:09
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 26 10:40:46 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 02:43:46 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.238	1.290	-4.2	99	0.00
89 C	Benzo(a)pyrene	1.174	1.163	0.9	95	0.00
90	Dibenzo(a,h)anthracene	0.907	0.965	-6.4	104	0.00
91	Benzo(a,h,i)perylene	0.887	0.956	-7.8	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0