


Data Path : Z:\HPCHEM1\BNA F\DATA\BF022717\
 Data File : BF093212.D
 Acq On : 27 Feb 2017 18:44
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 17 17:31:55 2017



AutoFind: Scans 832, 833, 834; Background Corrected with Scan 828

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.7	52749	PASS
68	69	0.00	2	1.7	781	PASS
69	198	0.00	100	35.8	45251	PASS
70	69	0.00	2	0.6	277	PASS
127	198	10	80	45.8	57830	PASS
197	198	0.00	2	0.3	438	PASS
198	198	100	100	100.0	126363	PASS
199	198	5	9	6.4	8150	PASS
275	198	10	60	22.8	28854	PASS
365	198	1	100	2.1	2670	PASS
441	442	0.01	24	15.7	14245	PASS
442	198	50	100	72.0	90930	PASS
443	442	15	24	19.4	17685	PASS