

Data Path : Z:\HPCHEM1\BNA F\Data\BF022718\
 Data File : BF103256.D
 Acq On : 27 Feb 2018 12:08
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 27 16:50:42 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 02:43:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.698	0.698	0.0	101	0.00
3	Pyridine	1.865	1.972	-5.7	106	0.00
4	n-Nitrosodimethylamine	0.752	0.767	-2.0	104	0.00
5 S	2-Fluorophenol	1.322	1.303	1.4	101	0.00
6	Aniline	2.335	2.237	4.2	99	0.00
7 S	Phenol-d6	1.618	1.535	5.1	100	0.00
8	2-Chlorophenol	1.374	1.336	2.8	100	0.00
9	Benzaldehyde	1.014	0.934	7.9	97	0.00
10 C	Phenol	1.878	1.783	5.1	99	0.00
11	bis(2-Chloroethyl)ether	1.426	1.415	0.8	102	0.00
12	1,3-Dichlorobenzene	1.570	1.529	2.6	101	0.00
13 C	1,4-Dichlorobenzene	1.574	1.512	3.9	100	0.00
14	1,2-Dichlorobenzene	1.451	1.380	4.9	100	0.00
15	Benzyl Alcohol	1.219	1.179	3.3	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.448	2.362	3.5	98	0.00
17	2-Methylphenol	1.248	1.217	2.5	101	0.00
18	Hexachloroethane	0.573	0.563	1.7	102	0.00
19 P	n-Nitroso-di-n-propylamine	1.007	0.932	7.4	100	0.00
20	3+4-Methylphenols	1.522	1.357	10.8	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.467	0.432	7.5	98	0.00
23 S	Nitrobenzene-d5	0.350	0.338	3.4	100	0.00
24	Nitrobenzene	0.386	0.372	3.6	99	0.00
25	Isophorone	0.690	0.658	4.6	100	0.00
26 C	2-Nitrophenol	0.182	0.189	-3.8	102	0.00
27	2,4-Dimethylphenol	0.277	0.262	5.4	97	0.00
28	bis(2-Chloroethoxy)methane	0.406	0.395	2.7	101	0.00
29 C	2,4-Dichlorophenol	0.266	0.259	2.6	100	0.00
30	1,2,4-Trichlorobenzene	0.283	0.270	4.6	98	0.00
31	Naphthalene	0.979	0.922	5.8	99	0.00
32	Benzoic acid	0.183	0.165	9.8	87	0.00
33	4-Chloroaniline	0.423	0.414	2.1	100	0.00
34 C	Hexachlorobutadiene	0.147	0.141	4.1	100	0.00
35	Caprolactam	0.093	0.094	-1.1	101	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.289	3.7	99	0.00
37	2-Methylnaphthalene	0.633	0.602	4.9	100	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
39	1,2,4,5-Tetrachlorobenzene	0.547	0.532	2.7	99	0.00
40 P	Hexachlorocyclopentadiene	0.152	0.130	14.5	82	0.00
41 S	2,4,6-Tribromophenol	0.169	0.160	5.3	93	0.00
42 C	2,4,6-Trichlorophenol	0.368	0.361	1.9	98	0.00
43	2,4,5-Trichlorophenol	0.423	0.402	5.0	95	0.00
44 S	2-Fluorobiphenyl	1.177	1.098	6.7	96	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF022718\
 Data File : BF103256.D
 Acq On : 27 Feb 2018 12:08
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 27 16:50:42 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 02:43:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.676	1.580	5.7	98	0.00
46	2-Chloronaphthalene	1.326	1.269	4.3	98	0.00
47	2-Nitroaniline	0.491	0.484	1.4	98	0.00
48	Acenaphthylene	2.040	1.919	5.9	97	0.00
49	Dimethylphthalate	1.604	1.555	3.1	100	0.00
50	2,6-Dinitrotoluene	0.337	0.330	2.1	98	0.00
51 C	Acenaphthene	1.190	1.110	6.7	97	0.00
52	3-Nitroaniline	0.427	0.423	0.9	99	0.00
53 P	2,4-Dinitrophenol	0.094	0.100	-6.4	102	0.00
54	Dibenzofuran	1.633	1.577	3.4	96	0.00
55 P	4-Nitrophenol	0.265	0.215	18.9	80	0.00
56	2,4-Dinitrotoluene	0.426	0.425	0.2	97	0.00
57	Fluorene	1.391	1.248	10.3	98	0.00
58	2,3,4,6-Tetrachlorophenol	0.284	0.268	5.6	94	0.00
59	Diethylphthalate	1.535	1.446	5.8	97	0.00
60	4-Chlorophenyl-phenylether	0.590	0.555	5.9	97	0.00
61	4-Nitroaniline	0.427	0.419	1.9	97	0.00
62	Azobenzene	1.562	1.468	6.0	96	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
64	4,6-Dinitro-2-methylphenol	0.112	0.117	-4.5	98	0.00
65 c	n-Nitrosodiphenylamine	0.696	0.662	4.9	97	0.00
66	4-Bromophenyl-phenylether	0.208	0.202	2.9	96	0.00
67	Hexachlorobenzene	0.226	0.220	2.7	99	0.00
68	Atrazine	0.209	0.195	6.7	95	0.00
69 C	Pentachlorophenol	0.109	0.094	13.8	83	0.00
70	Phenanthrene	1.101	1.019	7.4	95	0.00
71	Anthracene	1.129	1.053	6.7	95	0.00
72	Carbazole	1.124	1.055	6.1	96	0.00
73	Di-n-butylphthalate	1.406	1.329	5.5	96	0.00
74 C	Fluoranthene	1.106	1.025	7.3	95	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
76	Benzidine	0.748	0.708	5.3	92	0.00
77	Pyrene	1.559	1.573	-0.9	95	0.00
78 S	Terphenyl-d14	0.832	0.804	3.4	95	0.00
79	Butylbenzylphthalate	0.824	0.826	-0.2	92	0.00
80	Benzo(a)anthracene	1.298	1.273	1.9	91	0.00
81	3,3'-Dichlorobenzidine	0.463	0.440	5.0	86	0.00
82	Chrysene	1.260	1.214	3.7	90	0.00
83	Bis(2-ethylhexyl)phthalate	1.112	1.063	4.4	88	0.00
84 c	Di-n-octyl phthalate	1.796	1.702	5.2	88	0.00
85	Indeno(1,2,3-cd)pyrene	0.998	1.002	-0.4	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	92	0.00
87	Benzo(b)fluoranthene	1.315	1.327	-0.9	94	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF022718\
Data File : BF103256.D
Acq On : 27 Feb 2018 12:08
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 27 16:50:42 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 02:43:46 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.238	1.129	8.8	83	0.00
89 C	Benzo(a)pyrene	1.174	1.157	1.4	90	0.00
90	Dibenzo(a,h)anthracene	0.907	0.942	-3.9	97	0.00
91	Benzo(a,h,i)perylene	0.887	0.944	-6.4	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0