


Data Path : Z:\HPCHEM1\BNA F\Data\BF022718\
 Data File : BF103255.D
 Acq On : 27 Feb 2018 11:42
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF022218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Feb 27 17:05:20 2018



AutoFind: Scans 1623, 1624, 1625; Background Corrected with Scan 1619

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.0	52117	PASS
68	69	0.00	2	1.8	720	PASS
69	198	0.00	100	41.0	40290	PASS
70	69	0.00	2	0.8	336	PASS
127	198	10	80	54.1	53170	PASS
197	198	0.00	2	1.0	974	PASS
198	198	100	100	100.0	98296	PASS
199	198	5	9	6.6	6502	PASS
275	198	10	60	23.2	22757	PASS
365	198	1	100	2.7	2679	PASS
441	443	0.01	100	76.6	9862	PASS
442	198	50	100	67.3	66170	PASS
443	442	15	24	19.4	12869	PASS