


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022724\
Data File : BF137398.D
Acq On : 27 Feb 2024 11:20
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Feb 24 04:23:49 2024






AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1607

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.1	154760	PASS
68	69	0.00	2	1.5	2308	PASS
69	198	0.00	100	36.9	153871	PASS
70	69	0.00	2	0.5	764	PASS
127	198	10	80	43.0	179264	PASS
197	198	0.00	2	0.6	2294	PASS
198	198	100	100	100.0	416597	PASS
199	198	5	9	7.0	29029	PASS
275	198	10	60	26.8	111595	PASS
365	198	1	100	3.2	13138	PASS
441	198	0.01	100	16.9	70224	PASS
442	442	50	100	100.0	450560	PASS
443	442	15	24	19.0	85640	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022724\
Data File : BF137398.D
Acq On : 27 Feb 2024 11:20
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 27 12:06:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 24 04:23:49 2024
Response via : Initial Calibration



TIC: BF137398.D\data.ms

(70) Pentachlorophenol (C)

11.116min (-0.012) 931772.19 ng




response 426740

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.80	62.59
264.00	64.10	63.59
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022724\
Data File : BF137398.D
Acq On : 27 Feb 2024 11:20
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 27 12:06:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 24 04:23:49 2024
Response via : Initial Calibration



TIC: BF137398.D\data.ms

(77) Benzidine

12.674min (-0.006) 213243.78 ng m

response 2170448

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	17.10	14.29
183.00	11.00	11.68
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/27/2024	BNA_F	<u>BF137398.D</u>
Compound Name	Response	Retention Time
DDT	702972	13.492
DDD	9798	13.215
DDE	1089	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10887	713859	1.53

Instrument :
BNA_F
ClientSampleId :
DFTPP