




Data Path : Z:\HPCHEM1\BNA F\DATA\BF022817\
Data File : BF093238.D
Acq On : 28 Feb 2017 9:50
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 28 17:58:35 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 17:31:55 2017
Response via : Initial Calibration



TIC: BF093238.D

(69) Pentachlorophenol (C)




11.139min (-0.046) 36095.20ng

response 202066

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	63.69
264.00	62.90	64.65
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF022817\
 Data File : BF093238.D
 Acq On : 28 Feb 2017 9:50
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 28 17:58:35 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 17 17:31:55 2017
 Response via : Initial Calibration



TIC: BF093238.D

(76) Benzidine

12.671min (-0.046) 0.00ng

response 1437668

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	14.58
183.00	11.50	12.36
0.00	0.00	0.00