


Data Path : Z:\HPCHEM1\BNA_F\Data\BF022817\
 Data File : BF093238.D
 Acq On : 28 Feb 2017 9:50
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 17 17:31:55 2017



AutoFind: Scans 830, 831, 832; Background Corrected with Scan 827

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.7	48707	PASS
68	69	0.00	2	1.7	682	PASS
69	198	0.00	100	37.0	41178	PASS
70	69	0.00	2	0.5	196	PASS
127	198	10	80	44.4	49381	PASS
197	198	0.00	2	1.2	1334	PASS
198	198	100	100	100.0	111337	PASS
199	198	5	9	6.5	7245	PASS
275	198	10	60	23.2	25775	PASS
365	198	1	100	2.5	2739	PASS
441	442	0.01	24	15.5	15434	PASS
442	198	50	100	89.6	99768	PASS
443	442	15	24	19.3	19280	PASS