

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF022818\
 Data File : BF103307.D
 Acq On : 28 Feb 2018 21:38
 Operator : SJ/JU
 Sample : J1716-04 20X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-03-12-16

Manual Integrations
 APPROVED

Sohil
 3/1/2018 11:43:19 AM

Quant Time: Mar 01 01:09:16 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 17:05:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	151612	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	628091	20.00	ng	0.00
38) Acenaphthene-d10	9.91	164	255420	20.00	ng	0.00
63) Phenanthrene-d10	11.40	188	317575	20.00	ng	0.00
75) Chrysene-d12	14.05	240	235376	20.00	ng	0.00
86) Perylene-d12	15.55	264	243927	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	71297	7.11	ng	0.00
7) Phenol-d6	6.51	99	84950	6.93	ng	0.00
23) Nitrobenzene-d5	7.44	82	53401	4.86	ng	0.00
41) 2,4,6-Tribromophenol	10.71	330	8974	4.15	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	82244	2.04	ng	0.00
78) Terphenyl-d14	12.99	244	42065	4.30	ng	0.00
Target Compounds						
70) Phenanthrene	11.43	178	174463	9.982	ng	99
71) Anthracene	11.48	178	52075	2.906	ng	98
74) Fluoranthene	12.62	202	157408	8.963	ng	98
77) Pyrene	12.85	202	138134	7.527	ng	98
80) Benzo(a)anthracene	14.04	228	62696	4.105	ng	97
82) Chrysene	14.08	228	58629	3.954	ng	98
87) Benzo(b)fluoranthene	15.12	252	67783m	4.226	ng	
89) Benzo(a)pyrene	15.49	252	50040	3.494	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF022818\
Data File : BF103307.D
Acq On : 28 Feb 2018 21:38
Operator : SJ/JU
Sample : J1716-04 20X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-03-12-16

Manual Integrations
APPROVED

Sohil
3/1/2018 11:43:19 AM

Quant Time: Mar 01 01:09:16 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 17:05:20 2018
Response via : Initial Calibration

