



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022824\
Data File : BF137420.D
Acq On : 28 Feb 2024 13:55
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Feb 24 04:23:49 2024






AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1607

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.2	115915	PASS
68	69	0.00	2	1.3	1504	PASS
69	198	0.00	100	38.2	112864	PASS
70	69	0.00	2	0.5	610	PASS
127	198	10	80	44.6	131749	PASS
197	198	0.00	2	0.5	1484	PASS
198	198	100	100	100.0	295637	PASS
199	198	5	9	6.9	20545	PASS
275	198	10	60	24.9	73501	PASS
365	198	1	100	3.1	9222	PASS
441	198	0.01	100	14.8	43848	PASS
442	442	50	100	100.0	278101	PASS
443	442	15	24	19.7	54901	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022824\
Data File : BF137420.D
Acq On : 28 Feb 2024 13:55
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 28 14:45:28 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 24 04:23:49 2024
Response via : Initial Calibration



TIC: BF137420.D\data.ms

(70) Pentachlorophenol (C)

11.116min (-0.012) 829167.63 ng




response 309641

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.80	62.33
264.00	64.10	63.26
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022824\
Data File : BF137420.D
Acq On : 28 Feb 2024 13:55
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 28 14:45:28 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 24 04:23:49 2024
Response via : Initial Calibration



TIC: BF137420.D\data.ms

(77) Benzidine

12.674min (-0.006) 226416.11 ng

response 1897843

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	17.10	14.51
183.00	11.00	11.76
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/28/2024	BNA_F	<u>BF137420.D</u>
Compound Name	Response	Retention Time
DDT	510964	13.492
DDD	9005	13.192
DDE	1709	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10714	521678	2.05

Instrument :
BNA_F
ClientSampleId :
DFTPP