





Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF030116.M
Calibration Table Last Updated: Wed Mar 02 13:25:39 2016

Benzaldehyde


Response Ratio



$R = -8.42e-002 A^2 + 9.14e-001 A + 8.24e-002$
Coef of Det (r^2) = 0.992 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF030116.M
Calibration Table Last Updated: Wed Mar 02 13:25:39 2016


2,4-Dinitrophenol



$R = 2.49e-002 A^2 + 6.18e-002 A - 3.00e-002$
Coef of Det (r^2) = 0.997 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF030116.M
Calibration Table Last Updated: Wed Mar 02 13:25:39 2016


2,4-Dinitrophenol



$R = 2.49e-002 A^2 + 6.18e-002 A - 3.00e-002$
Coef of Det (r^2) = 0.997 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF030116.M
Calibration Table Last Updated: Wed Mar 02 13:25:39 2016


4,6-Dinitro-2-methylphenol



Resp Ratio = $1.41 \times 10^{-1} \times \text{Amt} - 5.09 \times 10^{-2}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF030116.M
Calibration Table Last Updated: Wed Mar 02 13:25:39 2016

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.41 \times 10^{-1} \times \text{Amt} - 5.09 \times 10^{-2}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF030116.M
Calibration Table Last Updated: Wed Mar 02 13:25:39 2016