

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF030118\
 Data File : BF103340.D
 Acq On : 1 Mar 2018 16:49
 Operator : SJ/JU
 Sample : J1720-19
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-42

Manual Integrations
 APPROVED

Sohil
 3/2/2018 3:30:43 PM

Quant Time: Mar 02 09:57:14 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 17:05:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	143551	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	575049	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	221794	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	318374	20.00	ng	0.00
75) Chrysene-d12	14.06	240	206290	20.00	ng	0.00
86) Perylene-d12	15.57	264	150335	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1064674	112.17	ng	0.01
7) Phenol-d6	6.52	99	1241165	106.87	ng	0.00
23) Nitrobenzene-d5	7.45	82	772823	76.75	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	166174	88.51	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	1017925	77.38	ng	0.00
78) Terphenyl-d14	12.99	244	672651	78.38	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.63	163	109018	6.127	ng	99
70) Phenanthrene	11.43	178	216637	12.364	ng	99
71) Anthracene	11.49	178	73383	4.084	ng	99
74) Fluoranthene	12.63	202	369720	20.999	ng	98
77) Pyrene	12.86	202	342529	21.297	ng	100
80) Benzo(a)anthracene	14.05	228	150982	11.280	ng	99
82) Chrysene	14.09	228	126320	9.720	ng	98
85) Indeno(1,2,3-cd)pyrene	17.10	276	48227	4.687	ng	98
87) Benzo(b)fluoranthene	15.13	252	134492m	13.606	ng	
88) Benzo(k)fluoranthene	15.15	252	35301m	3.793	ng	
89) Benzo(a)pyrene	15.50	252	88727	10.052	ng	# 92
91) Benzo(a,h,i)perylene	17.57	276	45306	6.797	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF030118\
Data File : BF103340.D
Acq On : 1 Mar 2018 16:49
Operator : SJ/JU
Sample : J1720-19
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-42

Manual Integrations
APPROVED

Sohil
3/2/2018 3:30:43 PM

Quant Time: Mar 02 09:57:14 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 17:05:20 2018
Response via : Initial Calibration

