

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF030118\
 Data File : BF103344.D
 Acq On : 1 Mar 2018 18:37
 Operator : SJ/JU
 Sample : J1721-13
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-46

Manual Integrations
 APPROVED

Sohil
 3/2/2018 3:30:51 PM

Quant Time: Mar 02 10:08:05 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 17:05:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	128790	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	492613	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	182430	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	282076	20.00	ng	0.00
75) Chrysene-d12	14.06	240	188108	20.00	ng	0.00
86) Perylene-d12	15.57	264	135952	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	810250	95.15	ng	0.01
7) Phenol-d6	6.52	99	936537	89.88	ng	0.00
23) Nitrobenzene-d5	7.44	82	578444	67.06	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	118848	76.97	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	794257	72.37	ng	0.00
78) Terphenyl-d14	13.00	244	606575	77.52	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.63	163	64830	4.430	ng	100
70) Phenanthrene	11.43	178	147632	9.510	ng	99
71) Anthracene	11.49	178	54180	3.404	ng	97
74) Fluoranthene	12.63	202	301959	19.357	ng	98
77) Pyrene	12.86	202	315831	21.535	ng	99
80) Benzo(a)anthracene	14.05	228	128011	10.488	ng	97
82) Chrysene	14.09	228	119999	10.126	ng	95
83) Bis(2-ethylhexyl)phthalate	14.02	149	22695	2.170	ng	97
85) Indeno(1,2,3-cd)pyrene	17.10	276	48241	5.142	ng	95
87) Benzo(b)fluoranthene	15.13	252	114707m	12.833	ng	
88) Benzo(k)fluoranthene	15.15	252	35512m	4.219	ng	
89) Benzo(a)pyrene	15.50	252	75438	9.451	ng	# 83
91) Benzo(a,h,i)perylene	17.57	276	44617	7.401	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF030118\
Data File : BF103344.D
Acq On : 1 Mar 2018 18:37
Operator : SJ/JU
Sample : J1721-13
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-46

Manual Integrations
APPROVED

Sohil
3/2/2018 3:30:51 PM

Quant Time: Mar 02 10:08:05 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 17:05:20 2018
Response via : Initial Calibration

