

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030215\
 Data File : BF077527.D
 Acq On : 2 Mar 2015 17:30
 Operator : TP/IZ
 Sample : G1399-11 2X
 Misc : S
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB5

Manual Integrations
 APPROVED

apatel
 3/3/2015 6:26:35 PM

Quant Time: Mar 03 05:50:49 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 02 23:46:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.26	152	105302	20.00	ng	0.00
21) Naphthalene-d8	8.84	136	420736	20.00	ng	0.00
38) Acenaphthene-d10	11.01	164	210121	20.00	ng	0.00
63) Phenanthrene-d10	12.85	188	415442	20.00	ng	0.00
75) Chrysene-d12	16.16	240	450027	20.00	ng	0.03
86) Perylene-d12	17.98	264	452257	20.00	ng	0.15
System Monitoring Compounds						
5) 2-Fluorophenol	5.55	112	236908	37.56	ng	0.00
7) Phenol-d6	6.82	99	325675	40.16	ng	0.00
23) Nitrobenzene-d5	7.95	82	192323	28.43	ng	0.00
41) 2,4,6-Tribromophenol	11.99	330	74323	36.74	ng	0.00
44) 2-Fluorobiphenyl	10.18	172	395666	29.36	ng	0.00
78) Terphenyl-d14	14.85	244	449669	23.36	ng	0.00
Target Compounds						Qvalue
48) Acenaphthylene	10.83	152	119131	6.03	ng	98
57) Fluorene	11.68	166	48736	3.35	ng	94
70) Phenanthrene	12.88	178	337102	14.00	ng	99
71) Anthracene	12.94	178	141795	5.93	ng	100
74) Fluoranthene	14.36	202	555856	21.13	ng	98
77) Pyrene	14.64	202	616053	22.38	ng	98
80) Benzo(a)anthracene	16.15	228	353095m	13.39	ng	
82) Chrysene	16.19	228	310644	12.54	ng	96
85) Indeno(1,2,3-cd)pyrene	19.45	276	189607m	6.71	ng	
87) Benzo(b)fluoranthene	17.51	252	408589m	15.54	ng	
88) Benzo(k)fluoranthene	17.54	252	122245m	4.74	ng	
89) Benzo(a)pyrene	17.91	252	300402m	11.99	ng	
91) Benzo(g,h,i)perylene	19.88	276	189259	7.85	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030215\
Data File : BF077527.D
Acq On : 2 Mar 2015 17:30
Operator : TP/IZ
Sample : G1399-11 2X
Misc : S
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB5

Manual Integrations
APPROVED

apatel
3/3/2015 6:26:35 PM

Quant Time: Mar 03 05:50:49 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 02 23:46:07 2015
Response via : Initial Calibration

