


Data Path : Z:\HPCHEM1\BNA F\Data\BF030217\
 Data File : BF093313.D
 Acq On : 2 Mar 2017 10:00
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 17 17:31:55 2017



AutoFind: Scans 831, 832, 833; Background Corrected with Scan 827

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.2	36659	PASS
68	69	0.00	2	1.9	608	PASS
69	198	0.00	100	37.7	31314	PASS
70	69	0.00	2	0.7	216	PASS
127	198	10	80	46.2	38365	PASS
197	198	0.00	2	1.2	992	PASS
198	198	100	100	100.0	83009	PASS
199	198	5	9	6.7	5544	PASS
275	198	10	60	22.3	18501	PASS
365	198	1	100	2.0	1642	PASS
441	442	0.01	24	16.4	9794	PASS
442	198	50	100	71.9	59670	PASS
443	442	15	24	19.2	11486	PASS