

Data Path : Z:\HPCHEM1\BNA F\DATA\BF030218\
 Data File : BF103372.D
 Acq On : 2 Mar 2018 18:25
 Operator : SJ/JU
 Sample : J1730-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ROLLOFF-20408

Manual Integrations
 APPROVED

Sohil
 3/5/2018 3:13:23 PM

Quant Time: Mar 02 22:46:33 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 02 13:28:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	136012	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	589930	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	264500	20.00	ng	0.00
63) Phenanthrene-d10	11.40	188	455010	20.00	ng	-0.01
75) Chrysene-d12	14.06	240	280985	20.00	ng	-0.01
86) Perylene-d12	15.56	264	233613	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	906543	100.81	ng	0.00
7) Phenol-d6	6.51	99	1091799	99.22	ng	0.00
23) Nitrobenzene-d5	7.44	82	721275	69.82	ng	-0.01
41) 2,4,6-Tribromophenol	10.71	330	207567	92.71	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	1046092	64.21	ng	0.00
78) Terphenyl-d14	12.99	244	860774	73.64	ng	0.00
Target Compounds						
10) Phenol	6.52	94	43950	3.440	ng	88
49) Dimethylphthalate	9.63	163	85185	4.015	ng	99
70) Phenanthrene	11.43	178	67484	2.695	ng	97
74) Fluoranthene	12.63	202	179687	7.141	ng	97
77) Pyrene	12.86	202	157551	7.192	ng	98
80) Benzo(a)anthracene	14.04	228	57762	3.168	ng	98
82) Chrysene	14.09	228	69983	3.953	ng	99
83) Bis(2-ethylhexyl)phthalate	14.02	149	463524	29.676	ng	# 99
85) Indeno(1,2,3-cd)pyrene	17.10	276	36488	2.604	ng	93
87) Benzo(b)fluoranthene	15.12	252	73658m	4.795	ng	
88) Benzo(k)fluoranthene	15.15	252	29502m	2.040	ng	
89) Benzo(a)pyrene	15.50	252	48498	3.536	ng	# 94
91) Benzo(a,h,i)perylene	17.57	276	38437	3.711	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF030218\
Data File : BF103372.D
Acq On : 2 Mar 2018 18:25
Operator : SJ/JU
Sample : J1730-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ROLLOFF-20408

Manual Integrations
APPROVED

Sohil
3/5/2018 3:13:23 PM

Quant Time: Mar 02 22:46:33 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 02 13:28:09 2018
Response via : Initial Calibration

