

Data Path : Z:\HPCHEM1\BNA F\DATA\BF030218\
 Data File : BF103384.D
 Acq On : 2 Mar 2018 23:41
 Operator : SJ/JU
 Sample : J1724-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-6

Manual Integrations
 APPROVED

Sohil
 3/5/2018 3:14:00 PM

Quant Time: Mar 05 03:59:41 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 02 13:28:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	144688	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	596041	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	220562	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	306292	20.00	ng	0.00
75) Chrysene-d12	14.07	240	219207	20.00	ng	0.00
86) Perylene-d12	15.58	264	155387	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1014649	106.06	ng	0.00
7) Phenol-d6	6.52	99	1252806	107.02	ng	0.00
23) Nitrobenzene-d5	7.45	82	794281	76.10	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	145807	78.10	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	1003593	76.53	ng	0.00
78) Terphenyl-d14	13.00	244	617622	67.73	ng	0.00
Target Compounds						
15) Benzyl Alcohol	7.03	79	36701	4.161	ng	Qvalue # 76
49) Dimethylphthalate	9.63	163	94551	5.344	ng	99
70) Phenanthrene	11.43	178	172065	10.208	ng	98
71) Anthracene	11.49	178	54891	3.176	ng	99
74) Fluoranthene	12.63	202	332917	19.654	ng	99
77) Pyrene	12.86	202	344369	20.149	ng	100
80) Benzo(a)anthracene	14.06	228	157085m	11.044	ng	
82) Chrysene	14.09	228	159886m	11.577	ng	
83) Bis(2-ethylhexyl)phthalate	14.02	149	29026	2.382	ng	99
85) Indeno(1,2,3-cd)pyrene	17.12	276	55619	5.087	ng	95
87) Benzo(b)fluoranthene	15.13	252	152033m	14.881	ng	
88) Benzo(k)fluoranthene	15.16	252	57511m	5.978	ng	
89) Benzo(a)pyrene	15.52	252	103517	11.346	ng	# 94
90) Dibenzo(a,h)anthracene	17.13	278	15241m	2.162	ng	
91) Benzo(a,h,i)perylene	17.60	276	54668	7.934	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF030218\
Data File : BF103384.D
Acq On : 2 Mar 2018 23:41
Operator : SJ/JU
Sample : J1724-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
SP-6

Manual Integrations
APPROVED

Sohil
3/5/2018 3:14:00 PM

Quant Time: Mar 05 03:59:41 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 02 13:28:09 2018
Response via : Initial Calibration

