

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030315\
 Data File : BF077553.D
 Acq On : 3 Mar 2015 18:48
 Operator : TP/IZ
 Sample : G1401-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-06-2-27-15

Quant Time: Mar 04 02:05:59 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 03 23:57:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.24	152	50449	20.00	ng	0.00
21) Naphthalene-d8	8.81	136	197761	20.00	ng	0.00
38) Acenaphthene-d10	10.99	164	99158	20.00	ng	0.00
63) Phenanthrene-d10	12.83	188	176057	20.00	ng	0.00
75) Chrysene-d12	16.13	240	204770	20.00	ng	-0.02
86) Perylene-d12	17.92	264	211535	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	211047	69.84	ng	0.00
7) Phenol-d6	6.80	99	162920	41.93	ng	0.00
23) Nitrobenzene-d5	7.93	82	387808	121.94	ng	0.00
41) 2,4,6-Tribromophenol	11.97	330	149072	156.13	ng	0.00
44) 2-Fluorobiphenyl	10.16	172	669210	105.23	ng	0.00
78) Terphenyl-d14	14.83	244	669361	76.42	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	10.67	163	31868	4.57	ng	# 99
51) Acenaphthene	11.03	154	29630	5.32	ng	96
54) Dibenzofuran	11.25	168	27837	3.24	ng	# 79
57) Fluorene	11.67	166	30809	4.48	ng	# 93
70) Phenanthrene	12.86	178	44859	4.40	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030315\
Data File : BF077553.D
Acq On : 3 Mar 2015 18:48
Operator : TP/IZ
Sample : G1401-07
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-06-2-27-15

Quant Time: Mar 04 02:05:59 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 03 23:57:38 2015
Response via : Initial Calibration

