


Data Path : Z:\HPCHEM1\BNA F\Data\BF030315\
Data File : BF077561.D
Acq On : 3 Mar 2015 22:36
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Mar 04 09:10:00 2015



AutoFind: Scans 1077, 1078, 1079; Background Corrected with Scan 1074

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	6821	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	35.0	6351	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.4	8606	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	18160	PASS
199	198	5	9	6.6	1206	PASS
275	198	10	60	24.2	4388	PASS
365	198	1	100	2.0	359	PASS
441	442	0.01	24	17.1	2571	PASS
442	198	50	100	82.7	15025	PASS
443	442	15	24	20.0	3008	PASS